

Pyranometer


ISO9060: 2018 Class A

Sub-category "Fast response" and "Spectrally flat"

MS-80

MS-80S

MS-80U



EKO

1. Index

1.	Index	1
2.	Important User Information	2
2-1.	Contact Information	2
2-2.	Warranty and Liability	2
2-3.	About Instruction Manual	3
2-4.	Environment	3
2-5.	ISO/IEC 17025	4
2-6.	CE Declaration	5
3.	Safety Information	7
3-1.	General Warnings	7
4.	Introduction	9
4-1.	About the Pyranometer Series	10
4-2.	Field applications	12
4-3.	Application to the standards	12
4-4.	Package Contents	13
5.	Getting Started	14
5-1.	Parts Name and Descriptions	14
5-2.	Setup	16
5-3.	Wiring	18
5-4.	Connections	19
5-5.	Measurements	30
6.	Maintenance & Troubleshooting	34
6-1.	Maintenance	34
6-2.	Calibration and Measurement Uncertainty	36
6-3.	Troubleshooting	37
7.	Specifications	39
7-1.	Specifications	39
7-2.	Dimensions	42
7-3.	Output Cables	44
7-4.	Accessories List	45
	APPENDIX	46
A-1.	Radiometric Terms	46
A-2.	Pyranometer Characteristics	47
A-3.	Software (Hibi)	48
A-4.	Communication Specifications (Modbus RTU)	62
A-5.	Communication specification (SDI-12)	74
A-6.	About recalibration	77
A-7.	Temperature sensor(10k Ω NTC, S/N: S19046150 or earlier)	78
A-8.	MS-80 RTD Temperature Conversion Table (Pt100 Class A)	79
A-9.	MS-80S Setting Report	80
A-10.	Surge protection	81

2. Important User Information

Thank you for using EKO Products

Reading this manual is recommended prior to installation and operation of the product. Keep this manual in a safe and handy place for whenever it is needed. For any questions, please contact us at below:

2-1. Contact Information

EKO INSTRUMENTS CO., LTD.

Asia, Oceania Region

www.eko-instruments.com	EKO INSTRUMENTS Co., Ltd.	Tel: +81 [3] 3469-6713
info@eko-instruments.com	1-21-8, Hatagaya, Shibuya-ku	Fax: +81 [3] 3469-6719
	Tokyo, 151-0072 Japan	

Europe, Middle East, Africa, South America Region

www.eko-instruments.com	EKO INSTRUMENTS Europe B.V.	Tel: +31 [0]70 3050117
sales-eu@eko-instruments.com	Lulofsstraat 55, Unit 28,	
	2521 AL, Den Haag, The Netherlands	

North America Region

www.eko-instruments.com	EKO INSTRUMENTS USA Inc.	Tel: +1 408-977-7751
sales-usa@eko-instruments.com	111 North Market Street, Suite 300	Fax: +1 408-977-7741
	San Jose, CA 95113 USA	

2-2. Warranty and Liability

For warranty terms and conditions, please contact EKO Instruments or your distributor for further details.

EKO guarantees that all products have been tested to ensure the instrument meets its published specifications. The product warranty is valid only if the product has been installed and used according to the instructions provided in this operating manual.

In case any manufacturing defect[s] occurs, the defected part[s] will be repaired or replaced under warranty; however, the warranty will not be applicable if:

- Any modification or repair has been done by other than EKO service personnel.
- The damage or defect is caused by disrespecting the specifications published on the Product Sheet or Manual.
- The bubble level fails or stops working (this part will not be serviced or replaced by EKO under warranty)
- There is discoloration of the pyranometer body, sun shield, and cable within a range that does not affect the function and performance of the product.

2-3. About Instruction Manual

Copy Rights Reserved by EKO INSTRUMENTS CO., LTD. Making copies of whole or parts of this document without permission from EKO is prohibited.

This manual was issued: Oct. 13, 2023
Version Number: 8

2-4. Environment

1. WEEE Directive



[Waste Electrical and Electronic Equipment]

Although this product is not subject to the WEEE Directive 2002/96/EC, please make sure that it should not be disposed of in a landfill or with municipal or household waste. For proper processing, collection and recycling, please contact a specialist collection site or facility.

Disposing of this product correctly will help save valuable resources and prevent any potential negative effects on human health and the environment, which could otherwise arise from inappropriate waste handling.

2. RoHS Directive

EKO Instruments has completed a comprehensive evaluation of its product range to ensure compliance with RoHS Directive 2011/65/EU+(EU)2015/863 regarding maximum concentration values for substances.

As a result, all products are manufactured using raw materials which concentration level are less than the hazardous substances specified in RoHS Directive 2011/65/EU+ (EU) 2015/863.

2-5. ISO/IEC 17025

EKO Instruments Co. Ltd. calibration laboratory is accredited by Perry Johnson Laboratory Accreditation, Inc. [PJLA] to perform pyranometer and pyr heliometer calibrations in accordance with the requirements of ISO/IEC17025, which are relevant to calibration and testing.

EKO is a manufacturer who can offer calibration service for pyranometers and pyr heliometers in-house. Based on the applied calibration methods EKO provides the best quality solar sensor calibrations compliant to the international standards defined by ISO/IEC17025 / 9847 [Indoor method] and ISO9059 [Outdoor method] [Certification: L13-94-R2 / www.pjilabs.com]

ISO/IEC17025 provides a globally accepted basis for laboratory accreditation that specifies the management and technical requirements. With calibrations performed at the EKO Instruments laboratory we enable our customers to:

- Clearly identify the applied calibration methods and precision
- Be traceable to the World Radiation Reference [WRR] through defined industrial standards:
 - ISO9846 Calibration of a pyranometer using a pyr heliometer
 - ISO9847 Calibration of field pyranometer by comparison to a reference pyranometer
 - ISO9059 Calibration of field pyr heliometers by comparison to a reference pyr heliometer
- Obtain repeatable and reliable calibration test results through consistent operations

Our clients will obtain highly reliable data by using an ISO/IEC17025 calibrated sensor. Our Accredited lab is regularly re-examined to ensure that they maintain their standards of technical expertise.

2-6. CE Declaration



IMPORTANT USER INFORMATION



DECLARATION OF CONFORMITY

We: EKO INSTRUMENTS CO., LTD
1-21-8 Hatagaya Shibuya-ku,
Tokyo 151-0072 JAPAN

Declare under our sole responsibility that the product:

Product Name : Pyranometer
Model No. : MS-80, MS-80U

To which this declaration relates is in conformity with the following
harmonized standards of other normative documents:

Harmonized standards:

EN 61326-1:2006 Class A [Emission]
EN 61326-1:2006 [Immunity]

Following the provisions of the directive:


EMC-directive: 89/336/EEC

Amendment to the above directive: 93/68/EEC

Date: April 4, 2019

Position of Authorized Signatory: General Manager of R & D Department

Name of Authorized Signatory: Shuji Yoshida

Signature of Authorized Signatory: 

**DECLARATION OF CONFORMITY**

We: EKO INSTRUMENTS CO., LTD
1-21-8 Hatagaya Shibuya-ku, Tokyo
151-0072 JAPAN

Declare under our sole responsibility that the product:

Product Name: Pyranometer
Model No.: MS-80S

To which this declaration relates is in conformity with the following harmonized standards of other normative documents:

Harmonized standards:

EN 61326-1:2013 Class A (Emission)
EN 61326-1:2013 (Immunity)

Following the provisions of the directive:

EMC-directive : 2014/30/EU

Date: Apr. 23,2020

Position of Authorized Signatory: Manager of Quality Assurance Div.

Name of Authorized Signatory: Minoru Kita

Signature of Authorized Signatory: 

3. Safety Information

EKO Products are designed and manufactured with consideration for safety; however, please make sure to read and understand this instruction manual thoroughly to be able to operate the instrument safely in the correct manner.



WARNING CAUTION

Attention to user; pay attention to the instructions given on the instruction manual with this sign.



HIGH VOLTAGE WARNING

High voltage is used; pay special attention to instructions given on this instruction manual with this sign to prevent electric leakage and/or electric shocks.



3-1. General Warnings

1. Setup

- The installation base or mast should have enough load capacity for the instrument to be mounted. Fix the pyranometer securely to the base or mast with bolts and nuts; otherwise, the instrument may drop due to a gale or an earthquake, which may lead to unexpected accidents.
- Make sure to install the instrument and cables in a suitable location, and avoid submersion in water.
- Insert the output cable into the connector port on the back of the sensor unit and tighten it all the way. Push the connector in, and check to make sure the screw is tight. If the connection is loose, water can enter the unit and cause it to malfunction.
- When connecting this product to a measuring instrument for measurement, the shield wire of the output cable shall be connected to signal ground (SG) for MS-80 and to ground earth for MS-80S. Otherwise noise may occur in the measurement data. In addition, the surge protection circuit inside the pyranometer will not operate properly if the shield wire is not connected to ground earth.
- Although this product is tested to meet EMC Directive compliance requirements, it may not fully satisfy its primary specification/performance when using this product near following locations where strong electromagnetic wave is generated. Please pay attention to the installation environment.
 - Outdoor: High voltage power line, power receiver/distribution facility, etc.
 - Indoor: Large-size chiller, large rotation device, microwave, etc.
- Do not use this product in the environment where corrosive gas, such as ammonia and sulfurous acid gas, are generated. It may cause malfunction.
- Do not install in the area that cause salt damages. It may cause malfunction by paint peeling off or corrosion. When installing in the area with risk of salt damages, make sure to take following measures:
 1. Wrap the connector with self-fusing tape
 2. Change the fixing screw to bolt screw made of aluminum
 3. Run the cables in a plastic or metal pipe treated with salt-resistant paint such as molten zinc plating
 4. Periodically clean.
- Do not use this instrument in the vacuum environment.
- For proper grounding use the original cable provided.

- If the cable and main unit are in risk for getting damaged by birds and small animals, protect the cable and the main unit by using:
 1. Reflective tape
 2. Repellent
 3. Cable duct
 4. Installing bird-spike
- When using the configurable 0 to 1V output from MS-80S, please use a precision resistor 100Ω. The settings can be changed with the 485 / USB conversion cable and dedicated software.

2. Handling

- Be careful when handling instruments with glass domes. Strong impact to this part may damage the glass and may cause injuries by broken glass parts.
- When carrying any MS-80 model with the sun screen attached, always hold the instrument from the bottom. Holding only the sun screen part may lead to dropping the sensor as it comes off from the sun screen.
- For voltmeters, data loggers, etc. to be measured by connecting to MS-80, MS-80U, use a product with an input resistance of 100MΩ or more.

3. Signal Cable [MS-80S]

- Make sure to ground the signal cable. When grounding is insufficient, it may cause not only measurement error due to noise, but also cause electric shock and leakage accidents.
- Check the voltage and types of specified power supply before connecting this instrument. When improper power supply is connected, it may cause malfunction and/or accident.
- Use this instrument with 0.5A fuse connected to the power supply line in series. Without connecting the fuse, it has risks of generating heat and fire due to large-current flowing by the power supply when internal damage of the electronics occurs.

4. About RS485 Modbus RTU connection [MS-80S]

- This product supports communication through the RS485 Modbus RTU.
- It is recommended to use the optional EKO converter cable when connecting MS-80S to a PC.
- Depending on the USB-RS485 converter type, an additional termination resistor (120Ω) and/or pull-up/pull-down resistor (680Ω) is required for proper communication.
- When connecting to a RS485 (Modbus RTU) master peripheral device, an additional termination resistor (120Ω) and/or pull-up/pull-down resistor (680Ω) is required for proper communication.

4. Introduction

EKO's new generation sensor, the MS-80 broke with the rules of traditional pyranometer architecture. The innovative design was inspired by the combination of latest technologies and state-of-the-art thermopile sensor, enabling a breakthrough in unprecedented low zero-offset behaviour, fast sensor response, high accuracy and over time measurement stability. MS-80 remains a class-leader for ISO 9060:2018 Class A solar sensors today. It offers several advantages over the competition to perform the most accurate irradiance measurements against low cost of ownership.

The compact sensor with single dome is immune to thermal offsets and integrates all value-added functions to reduce maintenance. MS-80 pyranometer no longer have a drying cartridge, silica gel inside will keep the sensor dry.

The MS-80S has integrated Smart electronics providing advanced measurement functions with different industrial signal outputs (**Modbus RTU, 4-20mA**, configurable **0-10mA** and **0-1V** with external precision shunt resistor 100Ω). Combined with on-board sensors for remote diagnostics of the sensor temperature, relative humidity and tilt position.

For MS-80S with product serial number 'S23037500' or earlier, it is possible to switch between Modbus RTU and **SDI-12** digital output signals.

For the MS-80 series a range of accessories is available (MV-01 ventilator and heater, cable extensions and albedo mounting plates).

MS-80 unique properties:

Long term stability Compared to our conventional pyranometer, the long-term stability is further enhanced with improved airtightness and the sensor architecture with low sensitivity degradation properties.

Fast detector response The MS-80 series is based on the latest thermopile technology, with a response time of <0.5sec @95% or <1sec @99% for all models (ISO 9060:2018 Class A, Sub-category "Fast response" and "Spectrally flat"). Such response time is suitable for measuring solar irradiance, which changes momentarily.

Excellent temperature coefficient MS-80 has a extremely low temperature error in wide temperature range compared to most conventional pyranometers, and it provides linear output against solar irradiance.

Lowest zero off-set Compared to any conventional pyranometers with double-domes, additional ventilation and heating system. MS-80 has the lowest off-set error designed intrinsically to the concept. The combination of the isolated thermopile detector architecture and optics keep the sensor in thermal balance within variable atmospheric conditions.

Warranty and re-calibration Unique in the market the MS-80 has 5 years warranty and 5 years recommended re-calibration period due to the great long term stability of the detector.

ISO 9060:2018 The international industry standard for pyranometer classification, was revised to the 2nd edition in 2018. Along with this revision, pyranometers are classified in order of the highest grade "Class A", "Class B", "Class C" and for Pyranometers that meet the response time and spectral selectivity criteria, "fast response" and "spectrally flat" are classified as a sub-category. MS-80, MS-80S and MS-80U are class A pyranometer that meets the subcategory of "fast response and spectrally flat".

4-1. About the Pyranometer Series

EKO offers three different MS-80 pyranometer models each with different features. With this wide range of sensors, EKO pyranometers can meet all possible application requirements, ranging from PV module efficiency measurements to climatology research and material durability testing.

All MS-80 sensors give excellent durability. The sensor is airtight and can be deployed with little maintenance since the desiccant is incorporated inside. Solar sensors are applied outdoors, hence the detector black surface, optical components and sensor mechanics are constantly exposed to solar radiation, temperature and pressure differences. UV radiation known as harmful radiation to materials can change the chemical properties of substances irreversibly. In the case of the MS-80, the detector is totally isolated below the sensor optics surface, which is sealed, and can't be affected by a high dose of UV, moisture or pressure differences.

During production and inspection, the directional response and temperature dependency are measured and validated through a measurement report that comes with each sensor. Besides, EKO provides a unique calibration service for pyranometers compliant to the international standards defined by ISO/IEC17025 / 9847 [indoor method]. When an ISO/ IEC17025 calibrated sensor is purchased, EKO provides a calibrated sensor with a consistently low calibration uncertainty. The Accredited lab is regularly re-examined to ensure that they maintain their standards of technical expertise.

A setting report is provided with all parameters which were set and tested during manufacturing (current range settings, digital output communication settings).

In case of MS-80, MS-80S, with combination of optional MV-01 [ventilator + heater unit], reduces the dew condensation and accumulation of dusts and snow on glass dome by continuously blowing heated air. The ventilation unit with heater is recommended when the sensor is deployed with a chance of occurrence of condensation, snow and ice. IEC 61724-1 Class A monitoring systems should use the MS-80 series pyranometer with the MV-01 which has a minimum <2W power consumption.

In the following paragraphs, the three instrument types are described individually.

1. MS-80

The analog MS-80 pyranometer can be used as a reference sensor to measure the global broad-band solar radiation with a high accuracy thermopile sensor. With excellent temperature response and non-linearity characteristics, it provides optimal performance throughout the day.

Category of ISO9060: 2018

“Fast response and spectrally flat pyranometer of class A”

Key features:

- Response time [$<0.5s@95\%$, $<1s@99\%$]
- Excellent temperature response over a wide temperature range [-20 to $+50^{\circ}C$]
- Low offset characteristics
- 5 Year warranty and recalibration period

2. MS-80S

The MS-80S with smart sensor technology and onboard diagnostic functions. Different output types can be selected, which is a great benefit for system integrators who work with various industrial standards. This new Smart transducer will also have additional features such as internal temperature and humidity sensors and a tilt sensor for remote sensor diagnostics. These additional internal sensors will help the user to monitor the stability of the irradiance sensors as well as to ensure its proper installation and maintenance practices.

Up to 31 smart sensors per one Master can be connected in one network. The signal converter settings can be changed using the optional RS485 / USB converter cable and the configurator software.

Category of ISO9060: 2018

“Fast response and spectrally flat pyranometer of class A”

Key features:

- Response time [$<0.5s@95\%$, $<1s@99\%$]
- Excellent temperature response over a wide temperature range [-20 to $+50^{\circ}C$]
- Low offset characteristics
- 5 Year warranty and recalibration period
- Outputs (**Modbus RTU**, configurable **SDI-12**^(*), **4-20mA**, configurable **0-10mA** / **0-1V** with external 100 Ω precision shunt resistor)
- Low power consumption
- Wide voltage-supply input range [5 to 30 VDC]
- With built-in tilt / roll sensor to check the sensor position over time.
- Temperature and humidity sensors to monitor the inside temperature of the pyranometer and condition of the drying agent (silica gel) inside the pyranometer body.

(*) SDI-12 output is supported only for MS-80S with product serial number ' S23037500' or earlier.

3. MS-80U

MS-80U is the most compact and light-weight ISO 9060:2018 Class A pyranometer on the market. With a total weight of only 200 grams and low body height the sensor can be easily used for drone and UAV applications. The body is hermetically sealed and can be used in environments up to 15 km above sea level. The sensor temperature characteristics are optimized for the lower part of the temperature range.

Category of ISO9060:2018

“Fast response and spectrally flat pyranometer of class A”

Key features:

- Compact and light weight
- Response time [$<0.5s@95\%$, $<1s@99\%$]
- Excellent temperature response over a wide temperature range [-20 to $+50^{\circ}C$]
- Low offset characteristics
- 5 Year warranty and recalibration period

4-2. Field applications

MS-80 is made for accurate irradiance measurements in the field of solar energy and meteorological research. The MS-80 analog sensor is well suited to be used in traditional meteorological networks where analog sensors are still common. MS-80S with the Smart interface with multiple outputs is made for the solar energy industry to be used in different applications. For PV research, plant monitoring and SCADA (Supervisory Control and Data Acquisition) controlled systems. The MS-80S with digital output can be used in a multi sensor network for onsite monitoring of large PV plants. Digital sensor can be easily configured, With the MS-80S, users can connect using a standard laptop and 'Hibi'; a new, custom built programme developed by EKO to give users real-time access to the internal diagnostics, custom settings, and data on irradiance, humidity, internal temperature and tilt angle from the sensor. This data can also be acquired when the pyranometer is connected through the data logger digital serial communication port. Each pyranometer has its own unique communication address. This way multiple pyranometers can be connected in a network.

The MS-80S builds on the revolutionary qualities of the original MS-80 by adding a new internal diagnostics system, giving users remote visibility over body temperature, internal humidity, tilt and roll angle; helping them to ensure optimum performance without the need for regular physical checks. Coupled with the 5-year recalibration interval, the internal diagnostic capabilities of the MS-80S makes it an ideal solution for complex networks, hard to reach locations, and monitoring stations with restricted access.

4-3. Application to the standards

MS-80 models are qualified to be used within any application which refers to an international standard according to:

- ISO9060: 2018 Fast response and spectrally flat pyranometer class A (Pyranometer classification)
- IEC 61724-1:2017 Class A (PV monitoring system requirement)
- ASTM 2848-11 (Test Method for Reporting Photovoltaic Non-Concentrator System Performance)
- WMO-No. 8, seventh edition 2008 High quality pyranometer (Meteorological system requirement)
- ISO/TR 9901:1990 (Solar energy field pyranometers recommended practice for use)
- ASTM G183 - 05 (Standard Practice for Field Use of Pyranometers)

MS-80 models comply to the international standards for calibration and traceable to the WRR:

- ISO 9847 Calibration of field pyranometers by comparison to a reference pyranometer
- ISO 9846 Calibration of a pyranometer using a pyr heliometer
- ISO 17025 scope of accreditation

4-4. Package Contents

Check the package contents first; if any missing parts or any damage is noticed, please contact EKO immediately.

Table 4-1. Package Contents

Contents	MS-80	MS-80S	MS-80U
Pyranometer	○	○	○
Output Cable	○ ⁽¹⁾	○ ⁽¹⁾	○ ⁽²⁾
Sun screen	○	○	-
Instruction Manual	Not included in the package [Please download from EKO Website]		
Setting Report	-	○ ⁽³⁾	-
Calibration certificate	○		
Cosine response report	○		
Temperature dependency report	○		
Quick Start Guide	○	○	-
Fixing Bolts	[M5] x2 [Bolt Length: 75mm]		
Washers	[M5] x4		
Nuts	[M5] x2		

(1) In case of MS-80/80S, standard length is 10m for both signal/power cable. For different length of cables [e.g. to meet your application needs] please contact EKO or your local distributor.

(2) MS-80U cable length is 3m, no other optional lengths are available. The cable is not detachable from the body.

(3) Details about the setting report, see appendix A-9.

5. Getting Started

5-1. Parts Name and Descriptions

Each part name and its main function is described below.




Table 5-1. Parts Name [MS-80 / MS-80S]

Parts Name
A. Diffuser
B. Glass Dome
C. Detector
D. Body
E. Sun screen
F. Levelling Screw
G. Cable, Connector
H. Spirit Level
I. Knurling Screw
J. Detector temperature sensor
K. Internal temperature sensor

*Signal converter is built-in for MS-80S.

Figure 5-1. Pyranometer Parts Name [MS-80 / 80S]



Table 5-2. Parts Name [MS-80U]

Parts Name
A. Diffuser
B. Glass Dome
C. Detector
D. Body
E. Levelling Screw
F. Cable
G. Spirit Level

*The cable can't be removed from the pyranometer

Figure 5-2. Pyranometer Parts Name [MS-80U]

1. Glass Dome, Quartz Diffuser

A glass dome creates a sealed environment for the detector and protects it against dirt and rain. The dome of the EKO pyranometers is only transparent for radiation emitted by the sun. The combination of the glass dome and the Quartz diffuser give a 180 degrees field of view and cosine response.

2. Detector

The thermopile detector, which is the heart of the pyranometer generates a voltage signal that is proportional to the solar irradiance. The fast response detector has a low temperature dependency and excellent long term stability.

3. Sun screen, Body

MS-80 and MS-80S have a sun screen to prevent any excessive body temperature increase generated by direct sun light. Weather resistant metals are used to reduce the effects and degradation caused by heat radiation, rain, and windblown dusts.

4. Spirit level

The integrated spirit level is used for setup and maintaining the sensor in a horizontally levelled position.

* Please note: Depending on the operating environment, the bubble may eventually disappear from the level gauge as the liquid dissipates. The reference circle may also fade out, or cracks may occur.

5. Drying agent

Enclosed drying agents inside the sensor body keep the sensor inside dry, prevents condensation of humidity inside of the glass dome.

6. Cable and Cable Connector

MS-80/ 80S are shipped with a 10 meter long output cable as standard length*.

Durable materials are used for the cable and connector, and pin terminals are attached at the end of the cable for easy connection with a data logger terminal block.

* If longer cables, round terminals or fork terminals are required, please contact EKO Instruments or your distributor. [Also see [7. Specification, 7-4. Accessories List] for optional items.]

The connector should be protected by the self-fusing tape when the pyranometers are used in the areas with the risk of the salt damage.

MS-80U has a 3 meter long output cable already attached, and the leads of the cable are soldered.

5-2. Setup

In order to obtain high quality measurements from pyranometers, several criteria with respect to setup and mounting of the instruments have to be considered. Also see the Quick Start Guide for comprehensive setup instruction details [included in the package of MS-80, MS-80S].

The ideal mounting position for a pyranometer is a location which has a full hemispheric field-of-view without any obstructions [such as buildings, trees, and mountains]; however, it might be difficult to find such locations. Therefore, in practice, it is ideal to install in a position which is free from obstructions at 5° above horizon.

The setup location should be easily accessible for periodic maintenance [glass dome cleaning], and avoid surrounding towers, poles, walls or billboards with bright colors that can reflect solar radiation onto the pyranometer. A strong physical impact to the pyranometer can lead to product damage and/or may cause changes to the sensitivity.

1. Installing at Horizontal or Tilted Position

- 1) Check the installation base where the pyranometer has to be mounted and make sure it has two fixing holes with the appropriate pitch. The pitch sizes of the fixing holes are as follows [in mm]:

Table 5-3. Fixing Hole Pitch and Bolt Size for Pyranometers

	MS-80 / MS-80S / MS-80U
Fixing Hole Pitch	65 mm
Fixing Bolt Size	M5 x 75 mm

- 2) Remove a sun screen [except for MS-80U]
The sun screen can be removed by loosening the knurling screw and sliding it towards the spirit level direction.
*When carrying MS-80 or MS-80S with a sun screen attached, always hold the instrument from the bottom. Holding only the sun screen may lead to dropping the sensor as it comes off from the sun screen.
- 3) There are two ways to setup the pyranometer (Horizontal or inclined).

Installing at horizontal Position

Setup the pyranometer with the signal cable connector facing the nearest Earth's pole. In the Northern hemisphere, the connector should be orientated North, in the Southern hemisphere, the connector should be orientated South. Adjust the pyranometer in a horizontal position by using the 2 levelling screws observing the air bubble in the spirit level while manipulating the levelling screws. The instrument is levelled horizontally if the air bubble is in the centre ring. The sensitivity of the bubble vial is (0.1° / mm). If the pyranometer is not levelled properly, the pyranometer readings are affected by cosine and azimuth errors. Periodically check the spirit level and adjust the pyranometer's position if necessary.


Installing at an inclined angle

After the MS-80, MS-80S or MS-80U is adjusted to horizontal position in levelled surface, install it on tilted a tilted mount.



When installing the pyranometer, do not remove the levelling feet or fixed feet. If the levelling feet are removed, it may induce offsets to output values due to the thermal effects from the mounting plate.

In the case of MS-80S, it is possible to acquire tilt information obtained from the built-in tilt position sensor through Modbus RTU. The sensor position can be measured in two different planes (X,Y).



<Tilt (X axis)> (180° to 0) and (0 to -180°)
 <Roll (Y axis)> (180° to 0) and (0 to -180°)

Figure 5.3A Tilt angle (-180 to 0°) and (0 to 180°)
 Roll angle (-180 to 0°) and (0 to 180°)

The tilt and roll angles are relative to the calibrated X and Y axis of the sensor and surface mounting plane X' and Y'. When the sensor is positioned in the same plane X = X' and Y = Y', the measured tilt and roll angles (X and Y) correspond to plane of installation X' and Y'.

The pyranometer connector is an important reference for setup and corresponds to the X-axis reference plane.



For setup at horizontal plane: Setup the pyranometer with the signal cable connector facing the nearest Earth's pole. In the Northern hemisphere, the connector should be orientated North, in the Southern hemisphere, the connector should be orientated South. If the cable is not oriented to the nearest pole, it does not affect the angle measurement. The angle measurement reflects the tilt and roll of the corresponding sensor position.

For setup at inclined plane: Setup the pyranometer with signal cable aligned with the installation platform (X', Y'), see figure 5-3B. When the sensor is not aligned, the tilt (X) and roll (Y) measurement does not reflect the angle X' and Y', see figure 5-3C.




Figure 5-3B Tilt and roll angle corresponding to reference plane




Figure 5-3C Tilt and roll angle not corresponding to reference plane

- 4) Fasten the pyranometer to the base with the 2 bolts [included] and put the sun screen back on the pyranometer.

5-3. Wiring

To extend the cable lifetime, make sure that the cables are not exposed to direct sun light or rain/wind by lining the cable through a cable conduit. Cable vibrations will potentially cause noise in the output signal. Fasten the cable so that the cable does not swing or move by wind blowing. Exposure of the signal cable to excessive electromagnetic emissions can cause noise in the output signal as well. Therefore, the cable should be lined at a safe distance from a potential source generating EMC noise, such as an AC power supply, high voltage lines or telecom antenna.

1. Ground connection



<MS-80>

Please be sure to connect the shield wire of the cable to the SG of the measuring instrument side. If it is not connected, it may be affected by noise.

<MS-80S>

The pyranometer housing is connected to the shield wire of the cable, so the shield wire shall be connected to the ground on the measuring instrument side.

Do not connect the pyranometer housing to ground. It will be a two-point ground and it cause a noise.




Figure 5-4A. MS-80/80U ground connection method




Figure 5-4B. MS-80S ground connection method

2. Wiring procedure

Connect the output cable to the MS-80 / MS-80S: **Insert the output cable into the connector port on the back of the sensor unit, and twist clockwise to fasten it. Make sure to tighten the connector all the way.**



- 1) If the connection is loose, water can enter the unit and cause it to malfunction.
- 2) Secure the length of the cable to avoid it from being pulled loose.
- 3) Connect the power cable grounding wire to prevent electrical shocks.

5-4. Connections

The MS-80 and MS-80U have an analog output for irradiance and detector temperature measurements. The MS-80S is based on a digital Smart processing interface providing advanced measurement functions with different industrial signal outputs (**Modbus RTU, 4-20mA**, configurable **0-10mA** and **0-1V** with external precision shunt resistor 100Ω).

For MS-80S with product serial number 'S23037500' or earlier, it is possible to switch between Modbus RTU and **SDI-12** digital output signals.

Table 5-4. Pyranometer outputs

Output	MS-80	MS-80S		MS-80U
		Serial number 'S23037500' or earlier	Serial number 'S23037501' or later	
Analog (mV)	○	-	-	○
Modbus RTU (Default)	-	○	○	-
SDI-12	-	○	-	-
4 - 20 mA (Default)	-	○	○	-
0 - 10 mA	-	○	○	-
0 - 1V	-	○(*)	○(*)	-

(*) MS-80S 0-10mA output with external precision shunt resistor 100Ω

Table 5-5. Pyranometer output parameters

Output parameters	MS-80	MS-80S	MS-80U
Irradiance	○	○	○
Temperature	○	○(*)	○
Tilt position	-	○(*)	-
Alerts for abnormal internal humidity of the pyranometer	-	○(*)	-
Relative humidity	-	○(*)	-
Sensor information	-	○(*)	-

(*) MS-80S digital output

1. Analog output

The MS-80 and MS-80U have an analog output for irradiance and detector temperature.



- Use a measuring device with an input impedance of 100MΩ or more. Low input impedance may cause underestimation of the output voltage measured.
- Be sure to connect the cable shield to the frame ground of the measurement device, otherwise noises may occur.

- A. MS-80 (from S/N: S19046151, temperature sensor 3W Pt100 class A). Connect the 5 wires with corresponding color to the input terminals of the voltmeter or data logger (figure 5-5).




Figure 5-5. How to connect MS-80 (from S/N: S1904615)

[*] Always connect the shield cable to prevent signal noise.

[**] See the Appendix A-8 resistance to temperature conversion table [PT-100].

- B. MS-80 (Till S/N: S19046150, temperature sensor type NTC 10kΩ). Connect the 4 wires with corresponding color to the input terminals of the voltmeter or data logger (figure 5-6).




Figure 5-6. How to connect MS-80 (Till S/N: S19046150)

[*] Always connect the shield cable to prevent signal noise.

[**] See the Appendix A-7 resistor to temperature conversion table [NTC 10k Ω].

- C. MS-80U (Temperature sensor type NTC 10k Ω). Connect the 4 wires with corresponding color to the input terminals of the voltmeter or data logger (figure 5-7).




Figure 5-7. How to connect MS-80U

[*] Always connect the shield cable to prevent signal noise.

[**] See the Appendix A-7 resistor to temperature conversion table [NTC 10k Ω].

2. Analog output (4-20mA / 0-1V)

MS-80S has two analog outputs (4-20mA and 0-1V). The 4-20mA output is default, the 0-1V(0-10mA) output can be enabled through the “Hibi” software which can be downloaded from the EKO website (MS-80S product page). See appendix 3 to change the current output settings. The current output can be converted into a voltage output when a shunt resistor is used in series. In combination with an external 100Ω precision resistor (0.1%, 15 ppm) a 0-1V output can be made.

Current output (4-20mA)

Connect the 3 wires with corresponding color to the input terminals of the current meter or data logger (figure 5-8). Connect the power terminals to a DC power supply [8 to 30 V]. We recommend to use a fuse [0.5A] to the DC power supply line [+] for over current protection.



Interconnect the power supply (-) and 4-20mA / 0-10mA (-) line. The MS-80S current output requires a 3 wire connection.

- **To protect your equipment from noise and surges, the shield wire of the cable must be connected to ground at a single point. If it is not connected, the electronic equipment may be damaged.**
- **The shield wire of the cable is connected to the surge protection circuit inside the pyranometer itself and to the enclosure.**
- **Be sure to connect a fuse, as this may lead to a fire.**




Figure 5-8. How to connect MS-80S (4-20mA output)

Voltage output (0-1V)

The current output can be converted into a voltage output when a shunt resistor is used in series. The default output is 4-20mA. The 0-10mA (0-1V) output is disabled. The settings can be changed when the MS-80S is connected to a PC through the Hibi software, see appendix 3 for instruction.

Connect the 3 wires with corresponding color to the input terminals of the Voltmeter or data logger (figure 5-9). Connect the power terminals to a DC power supply [8 to 30 V]. We recommend to use a fuse [0.5A] to the DC power supply line [+] for over current protection. The precision shunt resistor (0.1%, 15 ppm) need to be placed across the voltage input terminals of the measurement device.



Interconnect the power supply (-) and 4-20mA / 0-10mA (-) line. The MS-80S voltage output requires a 3 wire connection.

- **To protect your equipment from noise and surges, the shield wire of the cable must be connected to ground at a single point. If it is not connected, the electronic equipment may be damaged.**
- **The shield wire of the cable is connected to the surge protection circuit inside the pyranometer itself and to the enclosure.**
- **Be sure to connect a fuse, as this may lead to a fire.**




Figure 5-9. How to connect MS-80S (0-1V output)



When a precision shunt resistor (0.1%, 15ppm) is used in combination with the 4-20mA output it is required to match the precision shunt resistor value with supply Voltage to compensate for the voltage drop across the shunt resistor. (See the Table 5-6.)

Table 5-6. Shunt resistor value at supply voltage

Supply voltage	Shunt resistor value
8VDC to 15VDC	100Ω to 250Ω
15VDC to 24VDC	250Ω to 500Ω
24VDC to 30VDC	500Ω

Data logger connection (Modbus RTU)

Connect the 4 wires of the sensor cable with the corresponding wire color to the RS485 communication port of the data logger master and power supply unit (figure 5-11). Connect the power terminals to a DC power supply [8 to 30 V]. We recommend to use a fuse [0.5A] to the DC power supply line [+] for over current protection.




Figure 5-11. How to connect MS-80S (Modbus RTU)

MS-80S can be configured within a multi sensor RS485 Modbus RTU communication network. Up to 31 sensors per one Master can be connected and be assigned with a unique address. For a multiple sensor network the sensors need to be configured according to the recommended RS485 configuration standards as shown in figure 5-12.

The master represents the data-logging device and slaves represent devices such as MS-80S or other serial devices in the same network. Connect the communication wires of the slave to the Modbus RTU communication input of the master. Connect a 120Ω termination resistor at the end of the communication line. The master device may have an integrated termination resistor and pull-up and pull-down resistors. If any communication issues occur, apply those separately.




Figure 5-12. Communication Connection with Modbus RTU



Apply a Termination resistor (typically 120 to 150Ω) : Typically reflections occur on long lines, resulting in a receiver misreading logic levels. Proper termination prevents reflections, improving data integrity.

Apply Pull-up and pull-down resistors (typically 680 to 850Ω): Necessary to keep the voltage level constant when the transmission line is in a high impedance state.



Communication errors may occur depending on the connection distance and the number of connections. In that case, please prepare and apply a RS485 booster or a repeater.

Communication problems may occur if Modbus RTU and SDI-12 are used at the same time.



- Although a surge protection board is embedded inside the pyranometer, if the cable length exceeds 10 m, an appropriate SPD or the like shall be added to protect the measurement system from lightning surges (see A-10).

Data logger connection (SDI-12)

For MS-80S with serial number 'S23037500' or earlier, it works in SDI-12 digital output mode by selection from the Hibi software.

Connect the 3 wires of the sensor cable with the corresponding wire color to the SDI-12 communication port of the data logger master and power supply unit (figure 5-13). Connect the power terminals to a DC power supply [12 V]. We recommend to use a fuse [0.5A] to the DC power supply line [+] for over current protection.



Interconnect the power supply (-) and SDI-12 (-) line.

- **Communication problems may occur if Modbus RTU and SDI-12 are used at the same time.**



- **Although a surge protection board is embedded inside the pyranometer, if the cable length exceeds 10 m, an appropriate SPD or the like shall be added to protect the measurement system from lightning surges (see A-10).**




Figure 5-13. How to connect MS-80S (SDI-12)

MS-80S can be configured within a multi sensor SDI-12 communication network. Up to 20 units of S-series sensors can be connected and be assigned with a unique address (*). For a multiple sensor network the sensors need to be configured according to the recommended SDI-12 configuration standards as shown in figure 5-14. Make sure the cable length between the connecting slave and master be less than 60m. The shield cable shall be connected to the ground, otherwise noises may occur.

(*) When connecting other SDI-12 communication devices, the upper limit may change.




Figure 5-14. Communication Connection with SDI-12

4. Overview wire assignments

Table 5-7. Wire Color Codes of MS-80 and MS-80U (Also see [7-3. Output Cables])


No.	Cable Color		Signal	
	MS-80	MS-80U	MS-80	MS-80U
1.	Brown	Red	mV [+]	mV [+]
2.	White	White	mV [-]	mV [-]
3.	Blue	Green	Pt100 [B]	NTC
4.	Black	Black	Pt100 [B]	NTC
5.	Gray	---	Pt100 [A]	---
Shield	Shield	Shield	SG	SG

Table 5-8. Color Codes of cable of MS-80S

No.	Wire Color	4-20mA	Modbus RTU	SDI-12 [*]	0-1V [**]
1.	Brown (+)	8 to 30VDC (+)	5VDC(+) or 8 to 30VDC (+)	12VDC (+)	8 to 30VDC (+)
2.	White (-)	4~20mA (-) / GND	Vcc GND / RS485 GND	Vcc GND	0~10mA (-) / 0-1V (-) / GND
3.	Blue (+)	---	RS485 (+)	SDI-12 Data (+)	---
4.	Black (-)	---	RS485 (-)	---	---
5.	Gray (+)	4~20mA (+)	---	---	0~10mA (+) / 0-1V (+)
Shield	Shield	FG	FG	FG	FG

[*] SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.

[**] When selecting 0-1V output, a precision resistor is required separately.



Each number corresponds to the number in Table 5-7 and Table 5-8. There is no corresponding figure for MS-80U because the cable can't be removed from MS-80U body.

Figure 5-15. Connector pin number of MS-80 and MS-80S

5-5. Measurements

1. Data logger settings

The output voltage is measured by a measuring device such as a voltmeter or a data logger. If solar irradiance is measured continuously, it is recommended to use a data logger, which has sufficient memory to record data over a longer period and functions to process the measurement parameters of the pyranometer. Depending on the sensor output which will be used, the data logger should be capable to measure the assigned output type.

For MS-80 with analog output, the data logger should have a voltage input range (0-20mV), resolution ($\sim 1\mu\text{V}$) and impedance ($> 100\text{M}\Omega$). For the MS-80S with current output, Modbus RTU and SDI-12^(*), the data logger should have one of those inputs. For MS-80S the input range, resolution and impedance are not critical, and considered as standard for any meteorological or industrial type data logger. For current measurement the measurement range should be selected according to the range settings of the MS-80S.

When configuring the data logger parameters, the sampling rate and data reduction methods can be defined right at the beginning of the data acquisition process. The response time that is given in the specifications of the EKO pyranometers states the amount of time, which is necessary to reach 95% of the final measurement value. It is also possible to define a 63.2% response [which is equal to $1-1/e$]. This time constant, represented by the symbol τ , is 3 times smaller than the values specified by EKO. The recommended^[1] sampling rate for pyranometers is smaller than τ . So, for EKO pyranometers, the sampling rates that have to be programmed in the data logger systems should not exceed the values as given in Table 7-1.

Performing averaging and/or integration of measurement data can be meaningful to, e.g., reduce the data volume or to meet application-specific requirement. Note that shorter sampling rates allow to use shorter averaging/integration times [example for MS-80: 10 Hz sampling rate, 1 minute averaging period]. It could also be meaningful to store not only average values, but to keep track of all statistical values during the averaging period, namely: average, integral, minimum and maximum values, and standard deviation. As a general recommendation, the averaging/integration period should be as short as possible, but long enough to reduce the data volume to store the processed data safely.

Recommendations for irradiance measurements are explained in ^[1]“Guide to Meteorological Instruments and Methods of Observation”, WMO reference document No. 8.

^(*) SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.

2. Pyranometer default settings

The MS-80 and MS-80U have an analog output for irradiance and detector temperature measurements. The MS-80S is based on a digital Smart processing interface providing advanced measurement functions with different industrial signal outputs.

Where to find information regarding the MS-80 defaults settings and important parameters:

Table 5-9. MS-80 / 80U where to find the settings

MS-80 / 80U	Defaults setting and parameters
Sensitivity figure	<ul style="list-style-type: none"> - Calibration certificate - Serial number label
Impedance	<ul style="list-style-type: none"> - Serial number label

Table 5-10. MS-80S where to find the settings

MS-80S	Defaults setting and parameters
Sensitivity figure	<ul style="list-style-type: none"> - Calibration certificate - Modbus register - SDI-12^[1] register
Modbus RTU (Default)	<ul style="list-style-type: none"> - Setting report (see appendix A-9 MS-80S setting report) - Modbus register <ul style="list-style-type: none"> o Address (last two digits of serial number, 100 if “00”) o Baud rate 19200 o Parity Even
SDI-12 ^[1]	<ul style="list-style-type: none"> - Setting report - SDI-12 register <ul style="list-style-type: none"> o Address 0 o Baud rate 1200 o Parity Even
4 - 20 mA (Default)	<ul style="list-style-type: none"> - Setting report <ul style="list-style-type: none"> o 4-20mA = 0 – 1600W/m2
0 - 10 mA	<ul style="list-style-type: none"> - Setting report <ul style="list-style-type: none"> o 0-10mA = 0 – 1600W/m2
0 – 1V	<ul style="list-style-type: none"> - Setting report <ul style="list-style-type: none"> o 0-1V = 0 – 1600W/m2

^[1] SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.

3. Irradiance conversion

For MS-80, MS-80U with analog mV output the global solar Irradiance [W/m²] can be determined by measuring the output voltage [mV] divided by the individual sensor sensitivity [$\mu\text{V}/\text{W}\cdot\text{m}^{-2}$]. The maximum output voltage can be calculated by multiplying the maximum solar irradiance with the calibration factor [e.g. when the sensitivity of the MS-80 pyranometer is about $10\mu\text{V}/\text{W}\cdot\text{m}^{-2}$ or $0.010\text{mV}/\text{W}\cdot\text{m}^{-2}$, the maximum output voltage is about $1,400\text{W}/\text{m}^2$ times $0.010\text{mV}/\text{W}\cdot\text{m}^{-2} = 14\text{mV}$].

Analog output

The solar irradiance in Watts per meter squared [W/m²] is obtained when the output voltage E [μV] is divided by the sensitivity of the pyranometer S [$\mu\text{V}/\text{W}\cdot\text{m}^{-2}$]. This calculation is expressed by the following formula:

$$I [\text{W}/\text{m}^2] = \frac{E [\mu\text{V}]}{S [\mu\text{V}/\text{W}\cdot\text{m}^{-2}]}$$

*The sensitivity S for the pyranometer is stated on the calibration certificate and the product label.

4-20mA output

If the measurement range can be selected on the data logger, select the measurement range, which can accurately measure the signal within a range of 4 to 20mA. The output for the MS-80S is set to be $1,600\text{W}/\text{m}^2$ at 20mA.

When the solar irradiance current value is A [mA], the solar irradiance I [W/m²] can be determined by the following formula:

$$I [\text{W}/\text{m}^2] = (A [\text{mA}] - 4) \times (1600 / 16)$$

Note: When the range setting is changed for example 4-20 mA = 0-2000 [W/m²]

$$I [\text{W}/\text{m}^2] = (A [\text{mA}] - 4) \times (2000 / 16)$$

0-1V output

If the measurement range can be selected on the data logger, select the measurement range, which can accurately measure the signal within a range of 0 to 1V. The output for the MS-80S is set to be $1,600\text{W}/\text{m}^2$ at 1V.

When the solar irradiance voltage value is V [V], the solar irradiance I [W/m²] can be determined by the following formula:

$$I [\text{W}/\text{m}^2] = V [\text{V}] \times 1600$$

MS-80S [Modbus RTU Output, SDI-12^(*) Output]

When using the digital output (Modbus or SDI-12^(*)) by default the irradiance conversion is performed on-board and is one of the measurement parameters within the data string.

^(*) SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.

4. Daily radiant energy

The total daily radiant energy in Joule per meter squared [J/m^2] is obtained by integrating the solar irradiance over time. To calculate the total daily radiant energy in Joule per meter square [J/m^2], multiply the averaged solar irradiance I [W/m^2] by the averaging interval period [s]. Then sum-up the total data number [n] of averaged data points in one day.

Its physical unit is expressed with [J/m^2] and can be calculated with $J = W \cdot S$

$$\text{DTI} = \sum_{k=1}^n I_k \times t^k$$

6. Maintenance & Troubleshooting

6-1. Maintenance

By using the MS-80 pyranometers accurate results can be obtained if the glass dome and the condition of the instrument are maintained properly. Furthermore, regular maintenance and scheduled re-calibrations can extend the lifetime of the pyranometers. However, environmental conditions, such as for instruments mounted near highly frequented traffic lanes or airports, may have a deteriorating effect on the materials. Therefore, proper maintenance is needed and has to be adapted to the local environmental conditions.

The following table describes the common maintenance tasks that should be performed on a regular basis:

Table 6-1. Maintenance Items1

Maintenance Item	Frequency	How To	Effect
Clean Glass Dome	Several times per week	Keep the glass dome clean by using demineralized water and wiping with a soft cloth dry and clean.	The irradiance measurement will be affected due to a change in transmittance.
Check Appearance Condition	Weekly	Check for cracks and scratches on the glass dome and body.	May lead to shade on the detector and enhanced soiling.
Check Spirit level	Weekly	Verify if the pyranometer is levelled by checking the bubble is in the center ring of spirit level. [When the pyranometer is setup in horizontal position].	An additional cosine/azimuth error will be introduced.
Check Cable Condition	Weekly	Verify if the cable connector is properly connected, tightened to the instrument, and how cable is lined; make sure the cable is not shaking from the wind.	A disconnected cable will cause sporadic reading errors or failure of operation. If the cable is damaged, it may lead to noise or electric shock.
Check Setup Base Condition	Weekly	Check if the instrument is tightened properly to the mounting base plate and the base plate and/or table is securely fastened in a proper condition.	Loose instruments and/or mounting plates can lead to damages of the instruments and/or injury.
Check the Sun screen[*]	Weekly, Before/ After Bad Weather	Verify if the sun screen is securely fixed on the body, and knurling screw is securely tightened.	May lead to damaging the instrument and/or lead to increasing measurement error due to temperature increase by sun screen coming off.
Recalibration	Every 5 Years	To maintain the best possible measurement accuracy, recalibration of the pyranometer is recommended. Contact EKO for more details and requests for a recalibration and maintenance service.	Due to natural aging of materials the detector sensitivity of the pyranometer can gradually change over time.

※except for MS-80U

Table 6-2. Maintenance Items2 (Advanced remote check can be done on MS-80S)

Maintenance Item	Frequency	How To	Effect
Data validity	Weekly	Check the daytime irradiance data and compare to previous days or adjacent pyranometers.	When large difference occurs operating problems or installation issues can be detected.
Presence of noise	Weekly	Check night-time irradiance values	Night-time offsets and sensor stability issues can be revealed.
Check the temperature Inside	Weekly	[MS-80S] Check the body temperature via digital signal output.	If the inside temperature becomes abnormally high, the life of the product will be shortened.
Check the effect of desiccant	Weekly	[MS-80S] Check the relative internal humidity alert status of the pyranometer via digital signal output.	The condition of the drying agent can slightly change over time. If the relative humidity become high, the glass dome might be fogging up.
Check the tilt angle	Weekly	[MS-80S] Check the tilt angle and roll angle via digital signal output.	Any change in tilt position after the installation can affect the measurements due to the cosine response of the sensor.

6-2. Calibration and Measurement Uncertainty

It is recommended to recalibrate MS-80/MS-80S/MS-80U once every 5 years in order to verify the good quality of the solar radiation measurements. In the following paragraph the calibration method of EKO pyranometers and their calibration uncertainty is explained. For further information about recalibration and maintenance procedures, please contact EKO or find on the EKO website [<http://eko-eu.com>].

EKO can offer calibration service for pyranometers and pyrhemometers in-house. Based on the applied calibration methods EKO provides the best quality solar sensor calibrations compliant to the international standards defined by ISO/IEC17025 / 9847 [Indoor method] and ISO9059 [Outdoor method] [Certification: L13-94-R2 / www.pjilabs.com]

1. Calibration Method

MS-80 is calibrated indoors according to the ISO 17025 / 9847 international standard against a 1000W/m² AAA class solar simulator radiation source and designated calibration facility.

Indoor Calibration Procedure

Both reference and field pyranometers are placed in horizontal position in the center of the light beam at equal distance to the solar simulator light source. The pyranometers are alternatively exposed to the solar simulator beam hence the output is measured from each pyranometer for a specified time. From the reference pyranometer output [mV] and known sensitivity [$\mu\text{V}/\text{W}/\text{m}^2$] the sensitivity figure [$\mu\text{V}/\text{W}/\text{m}^2$] of the field pyranometer based on 10 data samples can be calculated.

Measurement Uncertainty of Indoor Calibration

The calibration uncertainty of MS-80 can be kept below (0.7%) as stated in the scope of accreditation and stated on each calibration certificate. Calibration in the lab is performed at a constant ambient temperature and ultra stable light conditions. Hence the repeatability of the method and pyranometers under test can be well maintained. The pyranometer calibration uncertainty is determined with consideration of uncertainty of the reference pyranometer and maximum variation of incident light during the measurement of field pyranometer and reference pyranometer. The expanded calibration uncertainty of each calibrated MS-80 model, and its results are stated on the calibration certificate.

2. Calibration Traceability

The MS-80 reference pyranometer is traceable to the absolute cavity pyrhemometer maintained at EKO. The absolute cavity pyrhemometer is directly traceable to the WRR [World Radiometric Reference] which is the Primary Standard Group of Absolute Cavities maintained at PMOD (Davos, Switzerland). The logger system used for the calibration measurement is traceable to JEMIC [Japan Electric Meters Inspection Cooperation].

Every year the MS-80 reference pyranometer is calibrated against the absolute cavity pyrhemometer, together with 5 other units part of the internal reference group of MS-80 pyranometers according the Shade / unshade Method [Method for Calibrating Reference and Field Pyranometers (1995)] Bruce W Forgan]. During the IPC the EKO absolute cavity is calibrated against the WRR every 5 years. Every year the EKO absolute cavity scale is examined during the NPC at NREL (Boulder, USA).

6-3. Troubleshooting

Read the following in case of any sensor trouble. If any questions should remain, please contact EKO for further technical support.

Table 6-3. Troubleshooting in field

Failure		Action
There is no output.	MS-80 MS-80U	Make sure that the sensor cable is connected properly to the instrument. To verify the connection, measure the impedance of output cable [between the “+” and the “-” wires] and check if the measured impedance is within the proper range as shown in the specification table.
	MS-80S	Make sure that the sensor is properly connected, and type of power supply and voltage values are appropriate. Also check the communication settings [i.e. port, baud rate, converter ID] are appropriate.
Output value is too low		<p>The glass dome maybe soiled with rain or dust. Clean the glass dome with demineralized water and soft cloth.</p> <p>The output may decrease over time. Recalibrate periodically.</p>
Negative output signal during night-time.		Pyranometers generate an output signal, which is proportional to the temperature differences between the sensor’s so-called hot and cold junctions. Night-time offset can occur when the dome temperature will cool down below the temperature of the detector. A slight negative offset within the specification can be expected.
Unusual noise		<p>Check the shield connection and make sure it is connected securely.</p> <p>Make sure if the output cable is not shaking from the wind; take necessary measure by fixing or lining the cables through metal pipe.</p> <p>Check for any objects, which emit electromagnetic wave around the instrument and or cable.</p> <p>When using a data logger or a measuring device with <100MΩ input impedance for MS-80 and MS-80U, a data logger potentially does not measure the sensor output correctly; thus take following measures in composition:</p> <ol style="list-style-type: none"> 1. Use a measuring device with input impedance more than 100MΩ 2. Set up the integration time and stability time as long as possible. 3. Use moving average processing on the data 4. Loop two or more ferrite cores around the end of the cable. 5. Avoid ground loops. Isolate the pyranometer electrically from the installation base. The sensor enclosure and the adjustment foot are electrically isolated by anodizing, but the isolation can be further strengthened by inserting an insulation sheet between the pyranometer and the installation base, and by inserting a resin washer in the enclosure fixing screw. 6. Connect the shield cable of the cable to an appropriate ground. 7. Use a data logger with differential voltage input.

<p>4-20mA and 0-1V output outside range Example) 2mA output when using 4-20mA. The fluctuation of measured current is large.</p>	<p>If it is confirmed that the wiring is correct, connect the ground wire of this product to the negative side of the data logger voltage (current) measurement terminal.</p> <p>Note that the appropriate ground wire connection method for this product may differ depending on the type of data logger and measurement environment. Select an appropriate ground wire connection method for this product according to the type of data logger and measurement environment.</p>
--	---

Table 6-4. Troubleshooting Hibi (MS-80S)

Failure	Action
No connection with PC	<p>Make sure that the sensor is properly connected, and power supply and voltage levels are appropriate. Also check the communication settings [i.e. port, baud rate, converter ID] are correct.</p>
No communication with USB cable	<p>If no COM port shows up when Hibi software is started.</p> <ul style="list-style-type: none"> - Reconnect the USB converter and refresh to search new COM devices - Check for COM port settings in Windows settings - Install an appropriate driver
Settings changes not saved to pyranometer	<p>Make sure that the sensor is properly connected, and power supply and voltage levels are appropriate. Also check the communication settings [i.e. port, baud rate, converter ID] are correct.</p> <p>Power off/on the pyranometer, some of the setting are effective after power off/on.</p>

7. Specifications

7-1. Specifications

1. Pyranometer Specifications

The comparison table below, Table 7-1, shows typical values for the characteristic parameters of the EKO Pyranometers and the corresponding values of the ISO 9060: 2018 standard. The content of the characteristic item is partly changed from ISO 9060: 1990. Please also refer to "A-2. Pyranometer Characteristics List".

Table 7-1. Pyranometer specifications

Characteristics		MS-80 / MS-80U	MS-80S
	ISO9060: 2018	Class A	
	[ISO9060: 1990]	[Secondary Standard]	
	Spectrally flat	Compliant	Compliant
Fast response	Compliant	Compliant	Compliant
Response time 95% output	<10 sec	<0.5 sec	<0.5 sec
Response time 99% output	---	<1 sec	<1 sec
Zero off-set a -200W/m ²	±7 W/m ²	±1 W/m ²	±1 W/m ²
Zero off-set b 5K/hr	±2 W/m ²	±1 W/m ²	±1 W/m ²
Total zero off-set (c)	±10 W/m ²	±2 W/m ²	±2 W/m ²
Non-stability	±0.8 %/1yr	±0.5 %/5yrs	±0.5 %/5yrs
Nonlinearity	±0.5 %	±0.2 %	±0.2 %
Directional response	±10 W/m ²	±10 W/m ²	±10 W/m ²
Spectral error	±0.5 %	±0.2 %	±0.2 %
Spectral selectivity	±3 %	±3 %	±3 %
Temperature response -10 to +40°C	±1 %	±0.5 %	±0.5 %
Temperature response -20 to +50°C	---	±0.5%	±0.5 %
Tilt response	±0.5 %	±0.2 %	±0.2 %
Additional signal processing error	±2 W/m ²	---	±1 W/m ²

Table 7-2. Other Specification

Characteristics	MS-80	MS-80S	MS-80U
Field of View	2π [sr]		
Wavelength range	285 to 3,000nm		
^[1] Operating temperature	-40 to +80°C Accuracy guaranteed temperature range: -20 to +50°C		
^[2] Maximum Irradiance	4,000W/m ²		
Spirit level accuracy	0.1 °		
Tilt sensor accuracy	---	<±1°	---
Humidity sensor accuracy	---	Nominal: ±2%RH	---
Temperature sensor of PCB (tolerance)	---	±0.5°C	---
Detector temperature sensor	Till S/N S19046150 NTC10kΩ From S19046151 Pt100 Class A ^[3] ±0.5%	Pt100 Class A ^[4] ±0.1%	10kΩ NTC
Internal temperature sensor	---	±0.5°C	---
Environmental Protection [IP Code]	IP67 Equivalent [IEC60529]		
Weight	0.35kg	0.41kg	0.21kg
Body	Anodized		
Sensitivity	Approx. 10μV/W·m ²	Approx. 10μV/W·m ² [4-20mA : 0-1600W·m ²] ^[5] [0-1V : 0-1600W·m ²] ^[5]	Approx. 10μV/W·m ²
Impedance	Approx. 45kΩ	---	Approx. 45kΩ
Output Cable [Outer diameter]	AWG22 0.3mm ² x 5 pins [φ5.3-5.7mm]		AWG28 0.1mm ² x 4 pins [φ3.3mm]
Output Cable Terminal	Pin-Terminal [0.3-9.5]		Solder
Output [or Signal]	Voltage [mV]	Modbus RTU (default), configurable SDI-12 ^[7] , 4-20mA(default) ^{[5] [6]} , configurable 0-10mA / 0-1V ^{[5] [6]} with external 100Ω precision shunt resistor	Voltage [mV]
Resolution	---	< 0.01W·m ⁻² ^[4]	---
Alert signal	---	Alert is issued when the internal humidity of the pyranometer is abnormal.	---

Input Power Supply	---	Modbus RTU: DC5V±5% or DC8V to DC30V±10% 0-1V/4-20mA: DC8V to 30V±10% SDI-12 ^[7] : DC9.6V to 16V	---
--------------------	-----	---	-----

^[1] When the instrument is used in the ambient temperature exceeding the accuracy assurance temperature range, the measurement error may increase.

^[2] The operational maximum irradiance is defined as the maximum irradiance exposure level. Beyond this point damage may occur to the sensor.

^[3] When using MS-80, please refer to the product label and check whether it corresponds to NTC or PT100.

^[4] Temperature sensor is internally connected to Modbus electronics.

^[5] When 4-20mA is 0 to 1600W·m⁻², 0-10mA/0-1V : 0 to 1600W·m⁻².(default)

^[6] Sensor setting can be changed by connecting the sensor to a PC (Use the USB cable for MS-80S (option) and download the free configuration software from the EKO website.

^[7] SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.

Table 7-3. Power consumption supply Voltage specific

	5V DC	12V DC	24V DC	Remarks
During stand-by	75mW (Approx. 15mA)	90mW (Approx. 7.5mA)	110mW (Approx. 4.5mA)	-
Modbus RTU	170mW (Approx. 34mA)	180mW (Approx. 15mA)	190mW (Approx. 8mA)	Peak value during communication
4-20mA output	-	300mW (Approx. 30mA)	640mW (Approx. 27mA)	Apply the shunt resistor 250 Ω and 20mA output
0-1V output	-	300mW (Approx. 30mA)	640mW (Approx. 27mA)	Apply the shunt resistor 100 Ω and 10mA output

7-2. Dimensions

1. MS-80 / MS-80S

Table 7-4. Dimensions [MS-80 / MS-80S]

	MS-80 / MS-80S
A. Fixing Hole Pitch	65 mm
B. Body Height	73 mm
C. Levelling Screw Height	16 mm
D. Width [including Sun screen/Cover]	Φ96 mm
E. Overall Height [approx.]	101mm




Figure 7-1. External Dimensions (MS-80/80S)

2. MS-80U

Table 7-5. Dimensions [MS-80U]

	MS-80U
A. Fixing Hole Pitch	65 mm
B. Body Height	50 mm
C. Levelling Screw Height	16 mm
D. Width	Φ78 mm




Figure 7-2. Outer Dimensions [MS-80U]

7-3. Output Cables

See [5-2. Setup, 2. Wiring] for instruction

1. MS-80 / MS-80S




Figure 7-3. Output Cables [MS-80 / MS-80S]

2. EKO Converter Cable (Option)




Figure 7-4. EKO Converter cable [MS-80S]

3. MS-80U




Figure 7-5. Output Cables [MS-80U]

7-4. Accessories List

Table 7-6. Accessories List

Option Items	Description	Pyranometers
Output cable	Cable Length: 10m, 20m, 30m, 50m Terminals: Fork Terminals, Round Terminals, Pin Terminals	MS-80 / 80S
Ventilation Unit with Heater	MV-01 ventilator and heater	MS-80 / 80S
EKO Converter Cable	Converts from RS485 → USB for the communication with MS-80S and allows to connect to PC via USB terminal. Cable Length:5m	MS-80S

APPENDIX

A-1. Radiometric Terms

Table A-1. Definitions of Terms

Global Solar Irradiance, Global Horizontal Irradiance [GHI]	Hemispherical solar irradiance received by a horizontal plane surface, expressed in units of W/m ² or kW/m ² .
Direct Solar Irradiance, Direct Normal Irradiance [DNI]	Normal-incidence solar irradiance received over a small solid angle which includes the circumsolar irradiance, expressed in units of W/m ² or kW/m ² .
Diffuse Solar Irradiance, Diffuse Horizontal Irradiance [DHI]	Global horizontal irradiance without the direct solar irradiance, i.e. indirect irradiance of the scattered solar radiation [by air molecules, aerosol particles, clouds, etc.], expressed in units of W/m ² or kW/m ² .
Pyranometer	A radiometer designed to measure the hemispheric solar irradiance over the wavelength range of about 300 to 3,000nm.
Pyrheliometer	A radiometer designed to measure the direct solar irradiance over a certain solid angle including the circumsolar irradiance.
World Meteorological Organization (WMO)	It is a specialized organization of the International Union that carries out international standardization and coordination of meteorological work. WMO: Abbreviation for World Meteorological Organization
World Radiation Reference [WRR]	Radiometric reference instrument system which has an uncertainty of less than +/-0.3%, expressed in SI units. This reference is maintained by the World Meteorological Organization [WMO], and it has been issued since January 1, 1980
ISO9060	An ISO norm [International Standard]. The first edition was published in 1990, then the second edition was revised in 2018. Based on the performance of each characteristic, Pyranometer is classified into three classes A, B, and C, and specifications of "Spectrally flat radiometer" and "Fast response radiometer" are set as sub-categories. Pyrheliometer is classified into 4 classes of AA, A, B and C based on the performance of each characteristic, and specifications of "spectrally flat radiometer" and "Fast response radiometer" are set as sub-categories.

A-2. Pyranometer Characteristics

Table A-2. Pyranometer Characteristics [see also ISO-9060:2018, CIMO Guide, WMO No. 8, 2008]

Response Time	The time [seconds] of a pyranometer sensor to reach 95% of its final output signal. [ISO 9060: 2018 added] If the response time reach to 95% is less than 0.5 seconds, "fast response" is attached to the applicable class as a subcategory.
Zero Off-Set a	Response [dark-signal] to 200W/m ² net thermal radiation [ventilated]
Zero Off-Set b	Response [dark-signal] to 5K per hour change in ambient temperature
Total Off-set c	[ISO 9060: 2018 added] Total zero off-set including the effects a), b) and other sources
Non-Stability	Rate of change [%] of the pyranometer sensitivity per year.
Nonlinearity	Percentage deviation from the responsivity at 500W/m ² due to any change of irradiance within the range 100W/m ² to 1000W/m ² .
Directional Response	Also referred to as cosine error [W/m ²]; the range of errors caused by assuming that the normal incidence responsivity is valid for all directions when measuring, from any direction, a beam radiation whose normal incidence irradiance is 1000W/m ² [ISO 9060: 2018 modified] Include the zenith angle of zenith angle 90 ° or more.
Spectral error	[ISO 9060: 2018 added] Maximum spectral mismatch error of Pyranometer [%] with respect to spectral irradiance at AM 1.5 and AM 5 under multiple atmospheric conditions on fine weather against the reference standard spectral irradiance defined by IEC60904-3: 2016 Photovoltaic devices - "Measurement principles for terrestrial photovoltaic [PV] solar devices with reference spectral irradiance data."
Spectral selectivity	Percentage deviation of the product of spectral absorptance and spectral transmittance from the corresponding mean within the range 0.35µm to 1.5µm.
Spectrally flat Pyranometer	[ISO 9060: 2018 added] If the spectral selectivity is less than 3 %, "spectrally flat" is attached to the applicable class as a subcategory.
Temperature Response	[ISO 9060: 2018 modified] Percentage maximum output error due to any change of ambient temperature between -10 to 40°C against the output at 20°C.
Tilt Response	[ISO 9060: 2018 modified] Percentage deviation from the responsivity at 0 ° tilt [horizontal] due to change in tilt from 0 ° to 180 ° at 1000W/m ² .
Additional processing errors	[ISO 9060: 2018 added] Error generated when converting the analog output of Pyranometer or Pyrhemliometer into a digital signal with a signal converter etc.

A-3. Software (Hibi)

'Hibi' means 'day by day' in Japanese, and with this program, users will be able to visualize detection signals, set communication parameters, and rapidly troubleshoot any issues from day one. You can change the settings of the MS-80S using the Windows software downloadable from the EKO website (MS-80S product page) and the optional USB cable.

1. Software Installation

Follow the steps below to install "Hibi" configurator software.

- 1) Download the latest version of the "hibi.zip" file [Compressed file: Zip format] from MS-80S product page on the EKO website.
- 2) Unzip "hibi.zip" file and check that the "setup.exe" file exists.
- 3) Run the "setup.exe" file to install "Hibi" configurator software.

2. Hardware Preparation

After the software is installed, connect the MS-80S with a PC to change the setting.

- 1) After installing the software, connect the USB connector of "RS485/USB Converter cable" to the PC.
- 2) Connect the ends of the signal cable of MS-80S to the communication terminal of "RS485/USB converter".
- 3) Connect the power lines of the signal cable to the power supply (DC5V or 8 to 30V), and turn on the power supply. (Refer to Figure A3-1.)




Figure A3-1. Hardware preparation

- 4) If you use the optional "EKO Converter cable", you can install it simply by inserting the cable connector into the main unit. Power for the MS-80S can be supplied via USB. (See Figure A3-2).

* This conversion cable is for installation, so please avoid using it outdoors for a long time. Because the ground wire is not connected, surges can damage your PC during a lightning strike.




Figure A3-2. Using "EKO Converter cable"

3. Software

Start the “Hibi” software and make the required settings.

- 1) Start “Hibi” software. (Launch the software from “hibi.exe” located in the shortcuts or installation folder.)
- 2) After the software starts, the “Welcome” window appears.




Figure A3-3. Welcome window

- 3) Automatic connection

When using the “Auto-connect” function to connect MS-80S with the “Hibi” software, select the COM port that the RS485/USB converter cable is connected to.




Figure A3-4. Com port select

Click “Select” of “COM port:” on the startup screen, and select the COM port number assigned to the RS485 / USB conversion cable from the displayed COM port numbers.

If you can't find a selectable COM port, click the Refresh button and click Select again to try again.




Figure A3-5. Com port select

Check that the power supply of the main unit is "OFF". Then, click "CONNECT" button.

If the COM port is not recognized, press the refresh button and try again.

Power on the MS-80S within 10 seconds once the "Connecting your sensor" message appears.




Figure A3-6. Connecting screen

When the connection is completed, the window automatically switches to the "Dashboard" window. In the upper-right corner of the display changes to "Pyranometer Connected".




Figure A3-7. Dashboard

If the connection fails, the screen returns to the "Welcome" screen and "CONNECTION FAILED" is displayed.




Figure A3-8. CONNECTION FAILED

Make sure the connection settings are correct. To perform the "Auto-connection" again, turn off the power of the MS-80S, select "COM port", and then click "CONNECT".

For manual connections, click "View settings" at the bottom of "CONNECTION FAILED" to go to "Settings screen". (As for how to set and connect by "Manual connection", refer to the procedures (4) Manual connection) from the next section.

4) Manual connection

After turning on the power of the MS-80S, click "CONNECT" without selecting "COM port".

Clicking "view settings" at the bottom of the "CONNECTION FAILED" screen moves to the "Settings" screen. (The same screen appears even when the "Auto-connection" fails. To perform manual connection, set and connect using this procedure.)

After that, click the "CONNECT" button in Connection status.




Figure A3-9. Manual connection

When the connection is completed, "Connected" is displayed.

"XXXX Pyranometer Connected" is also shown in the upper-right corner of the display.




Figure A3-10. Disconnect appear

When the connection fails, the "Oops..." screen is displayed with the cause of the connection failure.

When connection fail appears

Check that the cable connection, COM port and the communication requirements are correct.

When changing the connection condition after successful connection, set "COM port", "Baud rate" and "Parity" in the "Serial communication" setting of the "COMMUNICATION" items and set "Modbus address [1-247]" and "Register Type" in the "Addresses" setting.

After entering the settings, click "APPLY CHANGES"

A confirmation window will be displayed.

Click "YES, SAVE" to make the settings for the MS-80S.

The settings in "Settings" can be saved / read / exported to CSV files.

Click the button with icons in the upper-right corner of "Settings" window.

- "Save" button: Save the settings.
- "Read" button: Read out the settings.
- "Output" button: Output the settings to CSV file




Figure A3-11. Oops screen




Figure A3-12. Apply changes screen




Figure A3-13. Settings screen

5) Setting items

1. COMMUNICATION setting items

Item Name	Setting contents	Setting range
Mode	Modbus RTU	Switch to Modbus RTU interface (Default)
	SDI-12 [*]	Switch to SDI-12 interface.
Serial communication	Auto search	OFF/ON: Manual/Auto-connect
	COM Port	COM1 to COM256
	Baud rate	2400 / 4800 / 9600 / 19200 / 38400 / 115200bps
	Parity	Even / Odd / None
Addresses	Modbus address [1-247]	1 to 247 (last 2 digits of serial number is default address, 100 when "00")
	Register Type	S-series / M-series/ SMP-series/ SRD-series
	SDI-12 [*] address	0 to 9 / A to Z / a to z
Connection status	---	Connected / Disconnect

[*] SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.




Figure A3-14. Communication setting screen

2. "SENSOR INFORMATION" setting items

Item Name	Setting contents	Setting range
Model information	Model	Model name registered in the MS-80S (cannot be changed)
	Serial number	Serial number registered to the MS-80S (cannot be changed)
	MFG. date	Date of manufacturing (cannot be changed)
	Firmware version	Firmware version of MS-80S (cannot be changed)
	Hardware version	Hardware version of MS-80S (cannot be changed)
Calibration	Calibration date	Calibration date and time: Registered at the time of calibration by the manufacturer, but can also be registered by the user.
	Sensitivity [$\mu\text{V}/\text{W}/\text{m}^2$]	Sensitivity constant: Registered at the time of manufacturer calibration, but can also be registered at the user side.




Figure A3-15. Sensor information settings

3. "OUTPUT" setting items

Item Name	Setting contents	Setting range
Analog output	Enable analog output	Analog output OFF/ON (selected with the radio button) Either 4-20mA or 0-1V power can be selected. * Default is 4-20mA.
	4-20mA	Current(mA)/Irradiance(W/m^2): Setting of minimum and maximum irradiance (W/m^2) corresponding to the output.
	0-1V	Voltage(V)/Irradiance(W/m^2): Setting of minimum and maximum irradiance (W/m^2) corresponding to the output.
	Resistance (ohm)	Load resistor value for calculating output current at 0-1V output




Figure A3-16. Analog settings screen

4. "TEMPERATURE" setting items

Item Name	Setting contents	Setting range
UNIT	Preferred unit (temperature unit setting)	°C -Celsius / °F -Fahrenheit / K-Kelvin




Figure A3-17. Sensor settings screen

5. "SENSOR POSITION" setting items

The sensor position can be measured in two different planes (X, Y).

<Tilt (X axis)> (180° to 0) and (0 to -180°)

<Roll (Y axis)> (180° to 0) and (0 to -180°)

The tilt and roll angles are relative to the calibrated X and Y axis of the sensor and surface mounting plane X' and Y'. When the sensor is positioned in the same plane $X = X'$ and $Y = Y'$, the measured angles X and Y represent the corresponding plane of installation X' and Y'.



Clicking "ZERO ADJUSTMENT" allows you to adjust Tilt (X axis) and Roll (Y axis) of Tilting items using the present tilt angle as 0°. The spirit vial can be used as a reference to readjust the tilt and roll position. The MS-80S X and Y position is factory calibrated.




Figure A3-18. Sensor Position setting

6. DATA FILES setting items

[Measurements]

<Save file to> Enter the name of the measurement data storage folder.

[Sensor settings]

<Save file to> Enter the name of the setting file storage folder.




Figure A3-19. Data Files setting

6) Measurement (Dashboard)




Figure A3-20. Dashboard screen

[Overview]

- Measurement data of irradiance, temperature, humidity, and tilt angle are read from MS-80S every 1 second and displayed.

<Irradiance> Irradiance measurement data (Unit: W/m²)

<Temperature> Internal temperature measurement data.

Unit: selected from °C (Celsius), °F (Fahrenheit) or °K (Kelvin). (°C: default)

<Humidity> Humidity measurement data (Unit: "%RH")

<Tilt Angle> Tilt (X axis angle) and Roll (Y axis angle) data

[Real-time measurement view]



Figure A3-21. Real-time measurement view

<Real time measurement-Irradiance>

- Display measured irradiance data in a graph (1-second update).
- Drag the graph to change the horizontal axis position and use the mouse wheel to zoom in or out on the desired area.
- The time range of the graph is 00:00:00 to 23:59:59.
- Clicking “Back to default” resets the change of the setting of the graph.

<Data Logging>

Logging function of measurement data

- “Sample time:” Set the logging interval for measurement data.
Setting range: 00:01~01:00 (1 sec to 1 min)
- “Average interval:” Set the logging interval for averaged measurement data.
Setting range: 00:01~01:00 (1 sec to 1 min)
- “Measure intervals:” Set the time interval for data logging.
By checking "Continue next days", it is also possible to keep setting after the next consecutive days.
- “Start at:” Set the start time for data logging.
- “Ends at:” Set the end time for data logging.
- * Setting across the day is not possible. The start time must be earlier than the end time. (Start time < End time)
- “LOG DATA buttons:” Click to start logging measurement data.
When "STOP" is clicked while measurement data is being logged, logging stops.

7) Contact (Contact)

When you click the “Contact” button, the contact page on [EKO Instruments] website appears in the Browser.




Figure A3-22. Contact page @ eko-instruments.com

8) Sensor internal humidity abnormality alert

Hibi has a function to monitor the relative internal humidity of the pyranometer at all times. When the internal relative humidity exceeds the standard value due to external failure or deterioration of drying agent over time, Hibi judges it to be abnormal and displays an alert. If a pyranometer is left as is after the alert, the possibility of a problem such as condensation inside the pyranometer glass dome will increase, so please contact EKO.





Figure A3-23. Sensor internal humidity abnormality alert

9) Recalibration recommendation notice

If Hibi detects that 5 years (1825 days) have passed since the initial calibration date, a recalibration recommendation notice will appear.

If the above notification is displayed, please contact EKO for recalibration.

* In the case of leap years, notifications will be displayed from the day before.



FigureA3-24. Recalibration recommendation notice

10) Exiting the software (Hibi)

To exit Hibi, click the "X" in the upper-right corner of the window.

A-4. Communication Specifications (Modbus RTU)

A4-1. Outlines

- This device is compatible with Modbus RTU*. ASCII modes are not supported.
(* This is an open protocol for serial communication developed by Modicon.)
- The initial setting of the communication interface is Modbus RTU mode. To use the SDI-12^(*) mode, change the settings from "Hibi" software. After changing the setting, turn the power off and then on again to operate with SDI-12^(*). To return to Modbus RTU mode again, change the setting from "Hibi" software.
- It corresponds to RS485, half-duplex and double-wire multi-drop networks.
(* SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.

The communication specifications are as follows.

Table A4-1. Modbus RTU communication protocol

Optional item	Remarks
Electrical specifications	EIA-485
Connection form	Multi-drop method (1 for Master, 31 for Slave, 32 for total)
Communications protocols	Modbus RTU *1 (Slave)
Communication speed baud rate	2400, 4800, 9600, 19200 , 38400, 115200 bps
Data length	8 Bit
Stop bit	1 bit/2 bit *2
Parity bit	None/Odd/ Even
Communication distance	Max.1000m (ideal value)
Error detecting system	CRC-16

*1: An open protocol developed by Modicon for serial communication.

*2: The value of the stop bit is determined automatically by the parity bit setting.

A4-2. Function code

- There is no distinction between Coil and Discrete Input addresses.
- There is no distinction between Holding Register and Input Register addresses.
- Coil, Discrete Input, Holding Register, Input Register addresses start from "0".

Table A4-2. Supported Function Codes

Function code (hexadecimal)	Function
0x01	Read Coils
0x02	Read Discrete Inputs
0x03	Read Holding Registers
0x04	Read Input Register
0x05	Write Single Coil
0x06	Write Single Register
0x10	Write Multiple Registers

A4-3. Communication specification

- The address range can be used from 1 to 247.
- The maximum frame size that can be communicated at a time is 256 bytes (maximum read registers: 125, maximum write registers: 123).
- The baud rate (bit/sec) can be selected from 2400, 4800, 9600, 19200, 38400 and 115200.
- Parity can be selected from “Even”, “Odd” or “None”.
- The bit width is 8 bits, the stop bit is 1 bit when parity is present, and 2 bits when parity is absent.
- If the address and communication settings are changed, the equipment must be restarted.

Table A4-3. Communication setting

Setting No.	Bit rate	Parity
0	2400	None
1	2400	Even
2	2400	Odd
3	4800	None
4	4800	Even
5	4800	Odd
6	9600	None
7	9600	Even
8	9600	Odd
9	19200	None
10 (default)	19200	Even
11	19200	Odd
12	38400	None
13	38400	Even
14	38400	Odd
15	115200	None
16	115200	Even
17	115200	Odd

A4-4. Cable connection

- RS485 communication ground and the power ground are the same.

Be sure to connect the ground of RS485 communication interface to the power ground.

A4-5. Data format

Table A4-4. Data format to use.

Format	Abstract
U16	Unsigned 16bit Integer
S16	Signed 16bit Integer
U32	Unsigned 32bit Integer
S32	Signed 32bit Integer
F32	IEEE754 32bit floating point format
Str	ASCII characters string

- The byte order for communication is big endian.
2byte values are sent in the order of H byte → L byte
4byte values are sent in the order of H word → L word
Character strings are sent in the order of 1 byte from the beginning.
Allocation of each format is shown in the table below.

Table A4-5. Assignment of 8/16/32 bit value

8bit	0x12	0x34	0x56	0x78
16bit	0x1234(MSW)		0x5678(LSW)	
32bit	0x12345678			

Table A4-6. Assignment of 8/16 bits and F32

8bit	0x41	0x45	0x85	0x1E
16bit	0x4145(MSW)		0x851E(LSW)	
F32	12.345			

Table A4-7. Assignment of 8/16bit and Str

8bit	0x41	0x42	0x43	0x44
16bit	0x4142(MSW)		0x4344(LSW)	
Str	'ABCD'			

- Unless otherwise stated, the 32-bit register is in the order of high word (MSW) followed by low word (LSW).
The following table shows the cases where 0x12345678 is assigned to address n of Modbus register.

Table A4-8. Relationship between 32-bit values and Modbus registers

32-bit value	0x12345678
Modbus register (address n)	0x1234(MSW)
Modbus register (n+1)	0x5678(LSW)

A4-6. Outline of 1 Holding/Input register map

- The register map starts from address 0.
- The contents of registers No. 0 to 49 are switched according to the register type setting (No.103).
- Registers No. 0 to 99 are read-only.
- Registers after No.101 can be read or written to.

A4-7. No. 0-49 Register details

- The register map changes depending on the register type setting. The following table shows 4 register maps.
- See in parentheses for the unit.
- Fixed value if the content is numeric.
- The blank space is not used. It is filled with zero in U16 format.
- The abbreviations are as follows.

Addr: Modbus slave address (1-247)

Comp Irr: Irradiance value after compensation of temperature and linearity.

Raw.Irr: Irradiance before correction

AD: Input voltage

Sensitivity: Calibration factor

Cal. Date: Calibration date

Supply: Supply Voltage

Sensor Name: Sensor name (MS-80S)

Firm Ver: Firmware version number

Hard Ver: Hardware version number

L. Coeff: Linearity compensation factor, set by No.182-189

Table A4-9. No. 0-49 Register S-series type (default)

0	Model	U16	Model number of the transmitter. (MS-80S:0x0110)
1	0	U16	Fixed value
2	Comp.Irr	F32	Adjusted solar radiation intensity Unit: (W/m ²)
3			
4 to 7 are Reserve			
8	Pt100	F32	Sensor temperature Unit: (°C)
9			
10 to 13 are Reserve			
14	X-axis tilt angle	F32	X-axis component of the tilt angle Unit: (°)
15			
16	Y-axis tilt angle	F32	Y-axis component of the tilt angle Unit: (°)
17			
18	Raw.Irr	F32	Intensity of solar radiation before correction Unit: (W/m ²)
19			
20	ADmV	F32	Sensor output voltage Unit: (mV)
21			
22	Internal temperature	F32	Temperature measured by the internal temperature sensor Unit: (°C)
23			
24	Internal humidity	F32	Relative humidity measured by the internal humidity sensor Unit: (% RH)
25			
26	Alerts for abnormal internal humidity of the pyranometer	U32	Alerts for abnormalities in the internal humidity of the pyranometer Normal: 0, Abnormality occurs: 1
27			
28 to 49 are Reserve			

Table A4-10. No.0-49 register M-series type

0	Serial number	U16	Serial No. 1st /2nd ASCII
1			Serial No. 3rd/4th ASCII
2			Serial No. 5th/6th ASCII
3			Serial No. 7th/8th ASCII
4			Serial No. 9th/10th ASCII
5			Serial No. 11st/12nd ASCII
6			Serial No. 13rd/14th ASCII
7			Serial No. 15th/16th ASCII
8	Model number	U16	Model No. 1st /2nd ASCII
9			Model No. 3rd/4th ASCII
10			Model No. 5th/6th ASCII
11			Model No. 7th/8th ASCII
12			Model No. 9th/10th ASCII
13	Minimum solar radiation intensity	U16	0 (fixed)
14	Maximum solar radiation intensity	U16	1600 (fixed)
15	Measurement Mode	U16	2(fixed)
16	Sensitivity ($\mu\text{V}/\text{W}\cdot\text{m}^{-2}$)	F32	Sensitivity (LSW)
17			Sensitivity (MSW)
18	Temperature unit	U16	Unit: °C,1 (fixed)
19	AD(mV)	F32	Input Voltage (LSW)
20			Input voltage (MSW)
21	Comp.Irr(W/m^2)	F32	Corrected insolation intensity (LSW)
22			Corrected insolation intensity (MSW)
23	Temp(°C)	F32	Sensor temperature (LSW)
24			Sensor temperature (MSW)
25	Password	U16	8355 (fixed)
26 to 33 are Reserve			
34	L.Coeff. k1	F32	Linearity correction factor 1-k1 (LSW)
35			Linearity correction factor 1-k1 (MSW)
36	L.Coeff. k2	F32	Linearity correction factor 2-k2 (LSW)
37			Linearity correction factor 2-k2 (MSW)
38	L.Coeff. k3	F32	Linearity correction factor 3-k3 (LSW)
39			Linearity correction factor 3-k3 (MSW)
40	L.Coeff. k4	F32	Linearity correction factor 4-k4 (LSW)
41			Linearity correction factor 4-k4 (MSW)
42	Addr	U16	Node number
43	BPS	F32	Communication speed (LSW)
44			Communication speed (MSW)
45	Parity	U16	Parity bit
46 to 49 are Reserve			

Table A4-11. No. 0-49 register SMP-series type

0	65535	U16	Fixed value
1	100	U16	Fixed value
2	1	U16	Fixed value
3	0	U16	Fixed value
4	1	S16	Fixed value
5	Comp.Irr (W/m ²)	S16	
6	Raw.Irr (W/m ²)	S16	
7	0	S16	Fixed value
8	Temp (0.1°C)	S16	Sensor temperature
9	Supply (0.1V)	S16	
10	0	S16	Fixed value
11	0	S16	Fixed value
12	0	S16	Fixed value
13	0	S16	Fixed value
14	0	S16	Fixed value
15	0	S16	Fixed value
16	0	U16	Fixed value
17	0	U16	Fixed value
18	AD(0.01uV)	S32	
19			
20	0	S32	Fixed value
21			Fixed value
22	0	S32	Fixed value
23			Fixed value
24	0	S32	Fixed value
25			Fixed value
26	0	U16	Fixed value
27	0	U16	Fixed value
28	0	U16	Fixed value
29	0	U16	Fixed value
30	0	U16	Fixed value
31	0	U16	Fixed value
32	0	U16	Fixed value
33	0	U16	Fixed value
34	0	S16	Fixed value
35 to 40 are Reserve			
41	2020	U16	Fixed value
42	0	U16	Fixed value
43	Firm Ver	U16	
44	Hard ver.	U16	
45	Address	U16	
46 to 49 are Reserve			

Table A4-12. No. 0-49 Registers SRD-series type

0	Address	U16	
1	Com.Set	U16	
2	Comp.Irr(0.01W/m ²)	S32	
3			
4	Raw.Irr(0.01W/m ²)	S32	
5			
6	Temp (0.01°C)	S16	Sensor temperature
7	0	U16	Fixed value
8	100	U16	Fixed value
9	100	U16	Fixed value
10	AD(nV)	S32	
11			
12 to 31 are Reserve			
32	Sensor Name	Str	
33			
34			
35			
36			
37			
38			
39			
40	0	U16	Fixed value
41	Sensitivity ($\mu\text{V}/\text{W}\cdot\text{m}^{-2}$)	F32	
42			
43	0	U16	Fixed value
44	0	U16	Fixed value
45	0	U16	Fixed value
46	Cal.date (YYYYMMDD)	U32	
47			
48 to 49 are Reserve			

A4-8. Measurement value register update cycle

- The sensor output and the tilt angle measurement value in the register are automatically updated.

The update cycle is about 110 msec. If the data read cycle is less than 110msec, the same data will be read multiple times. The data read cycle should be 110msec or more.

A4-9. Register details from No.50 to 99

- A fixed value of zero (U16) has been written to an address that is not used.
- The "EKO" character is written as the company name registration in register No. 96 and 97.

Table A4-13. Register No.50-95

50 to 95 are Reserve

Table A4-14. Register No.96-99

96	Company name (0,1)	Str	The company name in ASCII format "EKO" is read.
97	Company name (2,3)		Four "EKO" characters, three "EKO" characters and one space (0x20).
98	Firmware version	U16	Firmware version number.
99	Hardware version	U16	Various parts of hardware

A4-10. No. 101 or later registers

- Registers after No.101 are readable and writable, and values written are immediately reflected.
- To save the written data, use Discrete Coil shown below to save the data.
- The 2 registers, Modbus address (No.101) and serial communication setting (No.102) related to communication are not reflected in the operation after saving and rebooting.

Table A4-15. Register No.100-199

100 is Reserve			
101	Modbus Address	U16	Modbus slave address
102	Serial communication setting	U16	Set bit rate and parity, see Table A4-3. "Communication setting"
103	Register Type	U16	No. Refer to the register type setting Table A4-9 - Table A4-12 of No.0-49 register.
104 to 105 are Reserve			
106	Setting the analog outputs	U16	Analog current output setting, Table A4-18. "Analog output setting details"
107 to 130 are Reserve			
131	0-1V loading resistor	F32	Accurate load resistance value, default 100 Ω
132			Used to calculate the output current in 0-1V mode.
133	Irradiance (0-1Vmode 0V)	F32	Sets the irradiance in 0-1V or 0V.
134			
135	Irradiance (0-1Vmode 1V)	F32	Sets the irradiance in 0-1V or 1V.
136			
137	Irradiance (4-20mode 4mA)	F32	Radiation intensity value at 4-20 mA mode and output current of 4mA defaults to 0W/m ²
138			
139	Irradiance (4-20mode 20mA)	F32	Radiation intensity value at 4-20mA mode and an output current of 20mA defaults use 1600W/m ²
140			
141 to 161 are Reserve			
162	Date of manufacture	U32	Date of manufacture in YYYYMMDD format
163			
164	Serial number	U32	Serial number with 32-bit integer value (maximum value 4294967295)
165			Enter the number part excluding the alphabet S.
166	Sensor name (0,1)	Str	Sensor name according to ASCII format, up to 16 characters Null-terminated unless all 16 characters are used
167	Sensor name (2,3)		
168	Sensor name (4,5)		
169	Sensor name (6,7)		
170	Sensor name (8,9)		
171	Sensor name (10,11)		
172	Sensor name (12,13)		
173	Sensor name (14,15)		
174 to 181 are Reserve			

182	L.Coeff.k1	F32	Linear correction factor k1
183			
184	L.Coeff.k2	F32	Linear correction factor k2
185			
186	L.Coeff.k3	F32	Linear correction factor k3
187			
188	L.Coeff.k4	F32	Linear correction factor k4
189			
190	Cal.Date	U32	Calibration date in YYYYMMDD format
191			
192	Cal.Value	F32	Calibration value, $\mu\text{V}/\text{W}\cdot\text{m}^2$
193			
194 to 199 are Reserve			

Table A4-16. Register No.200-219

200	Cal.Date history0	U32	Calibration History, Date 0
201			
202	Cal.Value history0	F32	Calibration history, sensitivity constant 0
203			
204	Cal.Date history1	U32	Calibration History, Date 1
205			
206	Cal.Value history1	F32	Calibration History, Sensitivity Constant 1
207			
208	Cal.Date history2	U32	Calibration History, Date 2
209			
210	Cal.Value history2	F32	Calibration History, Sensitivity Constant 2
211			
212	Cal.Date history3	U32	Calibration History, Date 3
213			
214	Cal.Value history3	F32	Calibration History, Sensitivity Constant 3
215			
216	Cal.Date history4	U32	Calibration History, Date 4
217			
218	Cal.Value history4	F32	Calibration History, Sensitivity Constant 4
219			

Table A4-17. Analog output setting details

Analog output setting value	Operation modes
0	Output OFF
1	0-1V Mode
2	4-20 mA Mode

A4-11. Linear correction factor

- Calculate the pyranometer output after linear correction using the following equation using the coefficient k1/k2/k3/k4 for the solar radiation meter output V(raw) after temperature correction.
- Linear post-correction sun photometer output $V(L) = k1 + (k2 \cdot V(\text{raw})) + (k3 \cdot V(\text{raw})^2) + (k4 \cdot V(\text{raw})^3)$
- The default value is k2=1, k1=k3=k4=0, so V(L)=V(raw), and the raw value is used for the corrected pyranometer output.

A4-12. Solar irradiance after correction

- After the linear correction and temperature correction written above, using the sensitivity value $S(\mu V/W \cdot m^2)$ for the output voltage, according to the following equation, calculate the irradiance value.

$$\text{Corrected solar irradiance: } I(C) = V(L) \times 1000/S$$

*Since the unit of V(L) is millivolts, multiply it by 1000 to obtain the irradiance value.

A4-13. Outline of Coil / Discrete Input Registers

- Reboot the device and save the settings by writing a specified bit.
- Do not write to any Coil other than Reboot/Save.
- The readout is always zero.

A4-14. Register details

Table A4-18. Coil register

Addr	Parameter	R/W	Type
0	-	-	-
1	Reboot	W	Bit
2	-	-	-
3	Save	W	Bit
4	-	-	-
5	-	-	-
6	-	-	-
7	-	-	-

A-5. Communication specification (SDI-12)

A5-1. Abstract

- SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.
- SDI-12 (Version:1.4) is supported.
- The initial setting of the communication interface is Modbus RTU. To use the SDI-12 interface, change the settings from "Hibi" software. After changing the setting, turn the power off and then on again to operate with SDI-12 mode. To return to Modbus RTU mode again, change the setting from "Hibi" software.
- When SDI-12 is enabled, the unit operates in "low power standby mode" with power consumption of approx. 2.5mA (at 12V) during non-communication.

A5-2. Communication specification

- The communication specifications of SDI-12 are as follows.

Table A5-1. SDI-12 Communication Specifications

Item	Description
Communications protocols	SDI-12 Version 1.4
Baud rate	1,200 bps
Data length	7bit
Stop bit	1
Parity bit	Even
Communication distance	Within 60 m

A5-3. Command list

- SDI-12 commands are listed below.
- The lowercase letter 'a' in the command list indicates SDI-12 address number.
- SDI-12 address number that can be set is 0-9, A-Z, a-z.

Table A5-2. SDI12 Commands

Command	Response	Description
?!	a<CR><LF>	Check the address number of the connected device. NOTE: If more than one device is connected, it will not function properly. Be sure to use the device alone.
a!	a<CR><LF>	Verify that the device with address number "a" is active.
aAb!	b<CR><LF>	Change the address number of the device with address number "a" to "b".
a!	Refer to the right section	Returns information. <Response example> "a14EKOINST_MS-80SV3220000001<CR><LF>" a: SDI-12 address 14: SDI-12 version (represents version 1.4) EKOInst: Company name (8 characters) MS-80S: Sensor model name (6 characters) V32: Sensor version (3 characters) 20000001: Serial number (8 characters)
aM!	a0001<CR><LF>	Request the device with address number "a" to execute measurement. 0001: indicates the 3-digit measurement execution time (000 indicates instantaneous), and the 4th digit indicates the number of data to be returned (1 piece).
aD0! aD1! aD2! aD3! aD4!	a+1000.0<CR><LF>	Request to send data to the device with address number "a". The output value is always appended with a sign (+ or -). If there is more than one output, the sign is also the delimiter. D0: Irradiance (W/m ²) 1 digit after decimal point D1: Sensor output voltage (mV) 4 digits after decimal point and Sensor temperature (Celsius) 2 digits after decimal point D2: X-axis tilt angle (degree) 1 digit after decimal point, forward/backward with connector facing back, positive value when back is up, Y-axis tilt angle (degree) 1 digit after decimal point, left/right with connector facing back, negative value when left is down D3: Internal temperature of pyranometer body (°C) 1 digit after the decimal point, Humidity inside the enclosure (RH %) 1 digit after the decimal point. D4: Alert to notify users of abnormality in the internal humidity of the pyranometer returns 0 for a normal condition and 1 for an abnormal condition. If the pre-measurement command is "MC", then three CRC characters are followed, please refer to SDI-12 standard for the content of CRC characters.

aMC!	a0011<CR><LF>	Request the device with address number "a" to start measurement and request the CRC to confirm that the command is correctly accepted. The response format is the same as "aM!"
aC!	a00101<CR><LF>	This device is the same as the "aM!" command except for the difference in the number of response characters. The character following a indicates the number of seconds (3 digits) and the number of data (2 digits).
aCC!	a00101<CR><LF>	This transmitter is the same as the "aMC!" command except for the difference in the number of response characters.
aR0!	a+0.0<CR><LF>	Request a measurement from the device with address number "a". Measurement is executed and a response is returned immediately. The answer is similar to "aD0!" to "aD2!" command.
aRC0!	a+0.0EmT<CR><LF>	Request measurement from the device with address number "a" and request error detection to verify that the command was accepted correctly. Measurement is executed and the measured value with three CRC characters added at the end is returned instantly.
aXSE!	a+XX.XX<CR><LF>	Read the sensitivity constant of this device. Two digits after the decimal point.
aXCD!	aYYYYMMDD<CR><LF>	Read the calibration date of this device. YYYY: AD, MM: month, DD: day.
METADATA	Variable depending on request data	Identify Measurement Commands and Identify Measurement Parameter Commands are supported.

A-6. About recalibration

When recalibrating at a recalibration center other than EKO, the calibration scale of the other center may differ slightly from the pyranometer calibration scale and sensitivity constants ($\mu\text{V}/\text{W}\cdot\text{m}^2$) obtained by EKO. The differences may be caused by differences in the calibration method, differences in the reference standard, measurement uncertainty, and measurement conditions.

If a new sensitivity constant was found and needs to be applied to the pyranometer, it can be applied in two ways:

- 1) (Post data correction) Multiply the relative difference of the irradiance value measured with the pyranometer. In this case, the original factory sensitivity value given by EKO remains unchanged. When the new sensitivity figure was found apply the ratio of the new and original calibration figure according to the formula below:

$$I_{new} [\text{W}/\text{m}^2] = \frac{E [\mu\text{V}]}{S_{original} [\mu\text{V}/\text{W}\cdot\text{m}^2]} \times \frac{S_{original} [\mu\text{V}/\text{W}\cdot\text{m}^2]}{S_{new} [\mu\text{V}/\text{W}\cdot\text{m}^2]}$$

Example:

It was proven that there was a relative difference in irradiance measured by MS-80S and the reference pyranometer. In this example the irradiance measured by MS-80S was lower than the irradiance measured by the reference pyranometer, a new sensitivity value can be calculated and applied by the following equation.

$$S_{new} = I_{MS80S} / I_{ref} \times S_{original}$$

S_{new}	New MS-80S Sensitivity ($\mu\text{V}/\text{W}/\text{m}^2$)
S_{origin}	MS-80S Original Sensitivity ($\mu\text{V}/\text{W}/\text{m}^2$)
I_{MS80S}	Irradiance measured by MS-80S (W/m^2)
I_{ref}	Irradiance measured by reference pyranometer (W/m^2)

- 2) (Change the sensitivity figure of the MS-80S) Set a new sensitivity figure to the MS-80S. This can be done via through the Hibi software. Connect the MS-80S to a PC with a RS485/USB converter cable. In settings, the new sensitivity figure can be applied. The original sensitivity figure remains, but will be replaced by the "last" sensitivity figure ($\mu\text{V}/\text{W}/\text{m}^2$). Multiple calibration figures can be stored inside the pyranometer, the "last" sensitivity figure is the actual one to calculate the irradiance.

A-7. Temperature sensor(10kΩNTC, S/N: S19046150 or earlier)

When determining the sensor temperature T (°C), convert the output R (resistance value, Ω) of the thermistor (44031 10kΩ NTC) using the following equation.

$$T = (\alpha + \beta (\text{LN}(R)) + \gamma (\text{LN}(R))^3)^{-1} - 273.15$$

T: Sensor Temperature (°C), R: Resistance Value (ohm), α : $1.0295 \cdot 10^{-3}$, β : $2.3910 \cdot 10^{-4}$, γ : $1.5680 \cdot 10^{-7}$

Table A7-1. Thermistor Temperature Conversion Table (44031, 10kΩ@25°C)

T (°C)	R (Ω)	T (°C)	R (Ω)	T (°C)	R (Ω)
-30	135200	0	29490	30	8194
-29	127900	1	28150	31	7880
-28	121100	2	26890	32	7579
-27	114600	3	25690	33	7291
-26	108600	4	24550	34	7016
-25	102900	5	23460	35	6752
-24	97490	6	22430	36	6500
-23	92430	7	21450	37	6258
-22	87660	8	20520	38	6026
-21	83160	9	19630	39	5805
-20	78910	10	18790	40	5592
-19	74910	11	17980	41	5389
-18	71130	12	17220	42	5193
-17	67570	13	16490	43	5006
-16	64200	14	15790	44	4827
-15	61020	15	15130	45	4655
-14	58010	16	14500	46	4489
-13	55170	17	13900	47	4331
-12	52480	18	13330	48	4179
-11	49940	19	12790	49	4033
-10	47540	20	12260	50	3893
-9	45270	21	11770	51	3758
-8	43110	22	11290	52	3629
-7	41070	23	10840	53	3504
-6	39140	24	10410	54	3385
-5	37310	25	10000	55	3270
-4	35570	26	9605	56	3160
-3	33930	27	9227	57	3054
-2	32370	28	8867	58	2952
-1	30890	29	8523	59	2854

A-8. MS-80 RTD Temperature Conversion Table (Pt100 Class A)

Table A8-1. Temperature conversion table [JIS C1604-2013 conformity] of Pt100(Class A)

T [°C]	R [Ω]	T [°C]	R [Ω]	T [°C]	R [Ω]
-30	88.2	0	100.0	30	111.7
-29	88.6	1	100.4	31	112.1
-28	89.0	2	100.8	32	112.4
-27	89.4	3	101.2	33	112.8
-26	89.8	4	101.6	34	113.2
-25	90.2	5	102.0	35	113.6
-24	90.6	6	102.3	36	114.0
-23	91.0	7	102.7	37	114.4
-22	91.4	8	103.1	38	114.8
-21	91.8	9	103.5	39	115.2
-20	92.2	10	103.9	40	115.5
-19	92.6	11	104.3	41	115.9
-18	92.9	12	104.7	42	116.3
-17	93.3	13	105.1	43	116.7
-16	93.7	14	105.5	44	117.1
-15	94.1	15	105.8	45	117.5
-14	94.5	16	106.2	46	117.9
-13	94.9	17	106.6	47	118.2
-12	95.3	18	107.0	48	118.6
-11	95.7	19	107.4	49	119.0
-10	96.1	20	107.8	50	119.4
-9	96.5	21	108.2	51	119.8
-8	96.9	22	108.6	52	120.2
-7	97.3	23	109.0	53	120.6
-6	97.7	24	109.3	54	120.9
-5	98.0	25	109.7	55	121.3
-4	98.4	26	110.1	56	121.7
-3	98.8	27	110.5	57	122.1
-2	99.2	28	110.9	58	122.5
-1	99.6	29	111.3	59	122.9

A-9. MS-80S Setting Report



EKO INSTRUMENTS CO.,LTD.
 1-21-8 Hatagaya, Shibuya-ku,
 Tokyo 151-0072 Japan
 P. +81.3.3469.6713
 F. +81.3.3469.6719
 www.eko.co.jp

S-series Setting Report(Example)

Issue Date: 10-Apr-21

Factory setting data	
Sensor type	Pyranometer
Model name	MS-80S
Serial number (s*****)	12345678
Manufacturing date	April 7, 2021
Calibration date	April 5, 2021
Analog sensitivity [μV/W/m2]	11.36
Hardware version	7
Firmware version (*.***)	3713

Current output mode of 4-20 [mA]	Setting
Enable/Disable	Enable (Default)
Output range corresponding to 4-20 [mA].	0 - 1600 [W/m2]

Current output mode of 0-10 [mA]	Setting
Enable/Disable	Disable (Default)
Output range corresponding to 0-10 [mA].	0 - 1600 [W/m2]

Modbus RTU	Setting
Enable/Disable	Enable (Default)
Address	1
Baudrate	19200
Parity	EVEN
Reg. type	S-series

SDI-12 (*)	Setting
Enable/Disable	Disable (Default)
Address	0
Baudrate	1200

(*) For MS-80S with product serial number 'S23037501' or later, there is no setting value listed for SDI-12.

A-10. Surge protection

Since a surge protection board is embedded inside the pyranometer, this alone provides strong protection against lightning surges.

However, if the pyranometer cable is longer than 10m, or if multiple pyranometer are connected and the total cable length exceeds 10m, the connected loggers, power supply, and other devices may be damaged by lightning surges.

Please refer to the figure below and an appropriate SPD shall be added to protect the measurement system.

A. Modbus RTU connection (when the cable length exceeds 10m)




Figure A10-1. Modbus RTU connection (when the cable length exceeds 10m, etc.)

*1 *2

If the cable length exceeds 10m, an SPD for the communication line and an SPD for the DC power supply shall be set up because surges may enter from the communication line and the power supply line.

B. SDI-12 (*) connection (when the cable length exceeds 10m)




Figure A10-2. SDI-12 connection (when the cable length exceeds 10m, etc.)

*1 *2

If the cable length exceeds 10m, an SPD for the communication line and an SPD for the DC power supply shall be set up because surges may enter from the communication line and the power supply line.

(*) SDI-12 output is supported only for MS-80S with product serial number 'S23037500' or earlier.

In the case of lightning surges, the more countermeasure components are used, the lower the probability of failure. However, the type, rating, and connection method of the SPD to be used vary depending on the manufacturer, so please read the instruction manual of each manufacturer carefully.



EKO Asia, Oceania

1-21-8 Hatagaya,
Shibuya-ku, Tokyo
151-0072 Japan
P. +81.3.3469.6711
F. +81.3.3469.6719
info@eko-instruments.com
www.eko-instruments.com

EKO North America

111 North Market Street,
Suite 300, San Jose,
CA 95113, USA
P. +1-408-977-7751
F. +1-408-977-7741
sales-usa@eko-instruments.com
www.eko-instruments.com

**EKO Europe,
Middle East, Africa,
South America**

Lulofsstraat 55, Unit 28,
2521 AL, Den Haag,
The Netherlands
P. +31 [0]70 3050117
sales-eu@eko-instruments.com
www.eko-instruments.com